



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108719

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/14/2020	17:20
2	CAL2	1	Water	NA	NA	20.2	7.01	04/14/2020	17:21
3	CAL3	1	Water	NA	NA	20.2	10.03	04/14/2020	17:22
4	ICV	1	Water	NA	NA	20.1	7.00	04/14/2020	17:25
5	CCV1	1	Water	NA	NA	20.1	2.02	04/14/2020	17:26
6	L2282-01	1	Water	NA	NA	18.8	11.77	04/14/2020	17:29
7	L2282-02	1	Water	NA	NA	19.0	3.86	04/14/2020	17:30
8	L2282-03	1	Water	NA	NA	19.2	3.73	04/14/2020	17:32
9	L2282-03DUP	1	Water	NA	NA	19.3	3.74	04/14/2020	17:33
10	CCV2	1	Water	NA	NA	20.1	12.02	04/14/2020	17:35

VB108719

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2282 WorkList ID : 137897 Department : Wet-Chemistry Date : 04-14-2020 17:10:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2020	Water	L2282-01	pH	Cool 4 deg C	PSEG01	H31	N48353	04/14/2020	9040C
04/14/2020	Water	L2282-02	pH	Cool 4 deg C	PSEG01	H31	1332	04/14/2020	9040C
04/14/2020	Water	L2282-03	pH	Cool 4 deg C	PSEG01	H31	FM5213	04/14/2020	9040C

Date/Time 04-14-2020 17:15
Received by: JO
Relinquished by: JO

Date/Time 04-14-2020 18:25
Received by: JO
Relinquished by: JO