



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108734

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/15/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	6.99	04/15/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	04/15/2020	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	04/15/2020	17:14
5	CCV1	1	Water	NA	NA	20.3	2.01	04/15/2020	18:05
6	L2279-07	1	Water	NA	NA	20.2	7.21	04/15/2020	18:06
7	L2298-02	1	Water	NA	NA	19.7	7.12	04/15/2020	18:08
8	L2298-02DUP	1	Water	NA	NA	19.8	7.13	04/15/2020	18:09
9	L2299-02	1	Water	NA	NA	20.6	4.81	04/15/2020	18:10
10	L2299-03	1	Water	NA	NA	20.6	5.05	04/15/2020	18:12
11	L2299-04	1	Water	NA	NA	20.3	8.18	04/15/2020	18:15
12	L2299-06	1	Water	NA	NA	19.7	7.82	04/15/2020	18:16
13	L2299-07	1	Water	NA	NA	19.5	7.95	04/15/2020	18:19
14	L2299-08	1	Water	NA	NA	19.5	7.79	04/15/2020	18:22
15	L2299-10	1	Water	NA	NA	20.4	4.71	04/15/2020	18:24
16	CCV2	1	Water	NA	NA	20.2	12.01	04/15/2020	18:25

108734

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2279 WorkList ID : 137920 Department : Wet-Chemistry Date : 04-15-2020 16:56:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2020	Water	L2279-07	pH	Cool 4 deg C	PSEG01	H31	OILY-WATER	04/14/2020	9040C
04/14/2020	Water	L2298-02	pH	Cool 4 deg C	SEAC01	G41	GROUNDWATER	04/14/2020	9040C
04/14/2020	Water	L2299-02	pH	Cool 4 deg C	PSEG01	H41	40860	04/14/2020	9040C
04/14/2020	Water	L2299-03	pH	Cool 4 deg C	PSEG01	H41	40861	04/14/2020	9040C
04/14/2020	Water	L2299-04	pH	Cool 4 deg C	PSEG01	H41	40862	04/14/2020	9040C
04/14/2020	Water	L2299-06	pH	Cool 4 deg C	PSEG01	H41	40863	04/14/2020	9040C
04/14/2020	Water	L2299-07	pH	Cool 4 deg C	PSEG01	H41	40864	04/14/2020	9040C
04/14/2020	Water	L2299-08	pH	Cool 4 deg C	PSEG01	H41	40865	04/14/2020	9040C
04/14/2020	Water	L2299-10	pH	Cool 4 deg C	PSEG01	H41	40511	04/14/2020	9040C

Date/Time 04.15.2020 16:56
 Received by: Ja
 Relinquished by: Ja

Date/Time 04.15.2020 18:30
 Received by: Ja
 Relinquished by: Ja