



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108735

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/15/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	6.99	04/15/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	04/15/2020	17:12
4	ICV	1	Water	NA	NA	20.2	7.00	04/15/2020	17:14
5	CCV1	1	Water	NA	NA	20.2	2.01	04/15/2020	18:05
6	L2279-08	1	Water	NA	NA	20.2	7.21	04/15/2020	18:06
7	L2298-02	1	Water	NA	NA	19.7	7.12	04/15/2020	18:08
8	L2298-02DUP	1	Water	NA	NA	19.8	7.13	04/15/2020	18:09
9	CCV2	1	Water	NA	NA	20.1	12.02	04/15/2020	18:25

VB 108735

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2279 WorkList ID : 137921 Department : Wet-Chemistry Date : 04-15-2020 16:56:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2020	Water	L2279-08	Corrosivity	Cool 4 deg C	PSEG01	H31	OILY-WATER	04/14/2020	9040C
04/14/2020	Water	L2298-02	Corrosivity	Cool 4 deg C	SEAC01	G41	GROUNDWATER	04/14/2020	9040C

Date/Time 04-15-2020 17:00
Received by: JS
Relinquished by: JS

Date/Time 04-15-2020 18:30
Received by: JS
Relinquished by: JS