


LB108741

7471B

INSTRUMENT ID : CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	70	0.000	70				
0.2	0.200	0.187	-0.013	1029	0.0 %	1029				
2.5	2.500	2.489	-0.011	13834	0.0 %	13834				
5.0	5.000	5.045	0.045	28049	0.0 %	28049				
7.5	7.500	7.439	-0.061	41368	0.0 %	41368				
10.0	10.000	10.026	0.026	55756	0.0 %	55756				

LB108741 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 16 Apr 2020 10:57:36							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50	70	-	0.00007471B	PPB		16 Apr 2020 11:00:26	S	Std		
0.2 - 1	50.2	1029	-	0.20007471B	PPB		16 Apr 2020 11:02:42	S	Std		
2.5 - 1	52.5	13834	-	2.50007471B	PPB		16 Apr 2020 11:04:59	S	Std		
5.0 - 1	55	28049	-	5.00007471B	PPB		16 Apr 2020 11:07:16	S	Std		
7.5 - 1	57.5	41368	-	7.50007471B	PPB		16 Apr 2020 11:09:32	S	Std		
10.0 - 1	510	55756	-	10.00007471B	PPB		16 Apr 2020 11:11:49	S	Std		
ICV99 - 1	ICV99	20923	3.7635	-7471B	PPB		16 Apr 2020 11:14:29	U	SMPL		
ICB99 - 1	ICB99	-37	-0.0047	-7471B	PPB		16 Apr 2020 11:16:44	U	SMPL		
CCV27 - 1	CCV27	27148	4.8826	-7471B	PPB		16 Apr 2020 11:18:59	U	SMPL		
CCB27 - 1	CCB27	-52	-0.0073	-7471B	PPB		16 Apr 2020 11:21:14	U	SMPL		
CRA - 1	CRA	1013	0.1841	-7471B	PPB		16 Apr 2020 11:23:29	U	SMPL		
HighStd - 1	HighStd	55278	9.9398	-7471B	PPB		16 Apr 2020 11:25:45	U	SMPL		
ChkStd - 1	ChkStd	39366	7.0791	-7471B	PPB		16 Apr 2020 11:28:00	U	SMPL		
PB128256BL - 1	PBS	-181	-0.0305	-7471B	PPB		16 Apr 2020 11:30:18	U	SMPL		
PB128256BS - 1	LCSS	19097	3.4352	-7471B	PPB		16 Apr 2020 11:40:15	U	SMPL		
L2272-01 - 1	WEST-SAMPLE	6852	1.2338	-7471B	PPB		16 Apr 2020 11:42:31	U	SMPL		
L2272-02 - 1	EAST-SAMPLE	19601	3.5258	-7471B	PPB		16 Apr 2020 11:44:47	U	SMPL		
L2279-01 - 1	OILY-STONE-POTHEAD	171	0.0327	-7471B	PPB		16 Apr 2020 11:47:04	U	SMPL		
L2279-03 - 1	OILY-STONE-REGULATORS	826	0.1505	-7471B	PPB		16 Apr 2020 11:49:22	U	SMPL		
L2283-01 - 1	PA-2	1493	0.2704	-7471B	PPB		16 Apr 2020 11:51:37	U	SMPL		
CCV28 - 1	CCV28	26776	4.8157	-7471B	PPB		16 Apr 2020 11:53:51	U	SMPL		
CCB28 - 1	CCB28	-21	-0.0018	-7471B	PPB		16 Apr 2020 11:56:06	U	SMPL		
L2283-02 - 1	PA-3	1711	0.3096	-7471B	PPB		16 Apr 2020 11:58:22	U	SMPL		
L2283-02DUP - 1	PA-3DUP	1537	0.2783	-7471B	PPB		16 Apr 2020 12:00:38	U	SMPL		
L2283-02MS - 1	PA-3MS	21882	3.9359	-7471B	PPB		16 Apr 2020 12:02:53	U	SMPL		
L2283-02MSD - 1	PA-3MSD	21186	3.8108	-7471B	PPB		16 Apr 2020 12:05:09	U	SMPL		
L2283-03 - 1	PA-1	1693	0.3064	-7471B	PPB		16 Apr 2020 12:07:25	U	SMPL		
L2283-04 - 1	PA-5	2237	0.4042	-7471B	PPB		16 Apr 2020 12:09:41	U	SMPL		
L2283-05 - 1	PA-4	1477	0.2675	-7471B	PPB		16 Apr 2020 12:11:56	U	SMPL		
L2283-06 - 1	WB-1-1	10921	1.9654	-7471B	PPB		16 Apr 2020 12:14:13	U	SMPL		
L2283-07 - 1	WB-1-2	6604	1.1893	-7471B	PPB		16 Apr 2020 12:16:30	U	SMPL		
L2284-02 - 1	TP-11-SA	7752	1.3956	-7471B	PPB		16 Apr 2020 12:18:45	U	SMPL		
CCV29 - 1	CCV29	28050	5.0448	-7471B	PPB		16 Apr 2020 12:21:01	U	SMPL		
CCB29 - 1	CCB29	-64	-0.0095	-7471B	PPB		16 Apr 2020 12:23:16	U	SMPL		
L2284-06 - 1	TP-11-DA	5482	0.9875	-7471B	PPB		16 Apr 2020 12:25:32	U	SMPL		
L2284-10 - 1	TP-12-SA	20767	3.7355	-7471B	PPB		16 Apr 2020 12:27:48	U	SMPL		
L2284-14 - 1	TP-12-DA	19369	3.4841	-7471B	PPB		16 Apr 2020 12:30:03	U	SMPL		
L2284-18 - 1	TP-16-B-SA	4355	0.7849	-7471B	PPB		16 Apr 2020 12:32:20	U	SMPL		
PB128266BL - 1	PBS	14	0.0045	-7471B	PPB		16 Apr 2020 12:34:36	U	SMPL		
PB128266BS - 1	LCSS	19727	3.5485	-7471B	PPB		16 Apr 2020 12:36:52	U	SMPL		
L2287-01 - 1	GLOUCESTER-SP	1378	0.2497	-7471B	PPB		16 Apr 2020 12:39:08	U	SMPL		
L2304-01 - 1	WC-B1	770	0.1404	-7471B	PPB		16 Apr 2020 12:41:26	U	SMPL		
L2304-01DUP - 1	WC-B1DUP	879	0.1600	-7471B	PPB		16 Apr 2020 12:46:29	U	SMPL		
L2304-01MS - 1	WC-B1MS	20911	3.7613	-7471B	PPB		16 Apr 2020 12:51:55	U	SMPL		
CCV30 - 1	CCV30	28118	5.0570	-7471B	PPB		16 Apr 2020 12:54:10	U	SMPL		
CCB30 - 1	CCB30	-71	-0.0108	-7471B	PPB		16 Apr 2020 12:56:25	U	SMPL		
L2304-01MSD - 1	WC-B1MSD	20831	3.7470	-7471B	PPB		16 Apr 2020 12:58:41	U	SMPL		
L2304-05 - 1	WC-B2	12	0.0042	-7471B	PPB		16 Apr 2020 13:00:57	U	SMPL		
L2304-09 - 1	WC-F1	864	0.1573	-7471B	PPB		16 Apr 2020 13:03:12	U	SMPL		
L2304-13 - 1	WC-F2	1295	0.2348	-7471B	PPB		16 Apr 2020 13:05:27	U	SMPL		
L2283-02LX5 - 1		339	0.0629	-7471B	PPB		16 Apr 2020 13:07:43	U	SMPL		
L2283-02A - 1		26258	4.7226	-7471B	PPB		16 Apr 2020 13:09:59	U	SMPL		
L2304-01LX5 - 1		153	0.0295	-7471B	PPB		16 Apr 2020 13:12:15	U	SMPL		
L2304-01A - 1		23931	4.3043	-7471B	PPB		16 Apr 2020 13:14:32	U	SMPL		
CCV31 - 1	CCV31	28127	5.0586	-7471B	PPB		16 Apr 2020 13:16:49	U	SMPL		
CCB31 - 1	CCB31	-62	-0.0091	-7471B	PPB		16 Apr 2020 13:19:05	U	SMPL		