



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108745

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/16/2020	11:25
2	CAL2	1	Water	NA	NA	20.2	7.01	04/16/2020	11:26
3	CAL3	1	Water	NA	NA	20.2	10.04	04/16/2020	11:28
4	ICV	1	Water	NA	NA	20.2	7.00	04/16/2020	11:30
5	CCV1	1	Water	NA	NA	20.2	2.02	04/16/2020	11:32
6	L2299-17	1	Solid	20.02	20	20.4	4.71	04/16/2020	11:35
7	L2306-01	1	Solid	20.04	20	20.8	11.30	04/16/2020	11:36
8	L2306-01DUP	1	Solid	20.05	20	20.8	11.31	04/16/2020	11:37
9	CCV2	1	Water	NA	NA	20.1	12.02	04/16/2020	11:40

VB 28108745

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2299

WorkList ID : 137947

Department : Wet-Chemistry

Date : 04-16-2020 11:13:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/14/2020	Solid	L2299-17	pH	Cool 4 deg C	PSEG01	H41	40511	04/14/2020	9045D
04/15/2020	Solid	L2306-01	pH	Cool 4 deg C	PSEG01	H42	CC-4-15-2020	04/15/2020	9045D

Date/Time 04-16-2020 11:20
 Received by: Jo
 Relinquished by: Jo

Date/Time 04-16-2020 12:25
 Received by: Jo
 Relinquished by: Jo