



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108746

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	04/16/2020	12:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/16/2020	12:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/16/2020	12:17
4	ICV	1	Water	NA	NA	20.1	7.00	04/16/2020	12:19
5	CCV1	1	Water	NA	NA	20.1	2.02	04/16/2020	12:20
6	L2315-01	1	Water	NA	NA	16.1	5.30	04/16/2020	12:22
7	L2315-01DUP	1	Water	NA	NA	16.2	5.31	04/16/2020	12:23
8	L2315-02	1	Water	NA	NA	17.0	4.76	04/16/2020	12:25
9	L2315-03	1	Water	NA	NA	18.0	4.89	04/16/2020	12:29
10	CCV2	1	Water	NA	NA	20.1	12.02	04/16/2020	12:30

2B108746

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2315 WorkList ID : 137952 Department : Wet-Chemistry Date : 04-16-2020 12:05:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/15/2020	Water	L2315-01	pH	Cool 4 deg C	STER09	G61	CELL-A-POOL-WATER-(DE-1K)	04/15/2020	SM 4500-H B
04/15/2020	Water	L2315-02	pH	Cool 4 deg C	STER09	G61	CELL-B-C-POOL-WATER-(DE	04/15/2020	SM 4500-H B
04/15/2020	Water	L2315-03	pH	Cool 4 deg C	STER09	G61	CELL-D-POOL-WATER-(DE-1K	04/15/2020	SM 4500-H B

Date/Time 04.16.2020 12:11:10
 Received by: Jm
 Relinquished by: AM

Date/Time 04.16.2020 13:20
 Received by: AM
 Relinquished by: Jo