



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/17/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 04/16/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:25
Out Date: 04/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108747

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L2251-01	METXB5	1	1.14	9.65	7.94	79.9	
L2251-02	METXB6	2	1.18	9.98	8.27	80.6	
L2251-03	METXB7	3	1.15	9.66	7.78	77.9	
L2251-04	METXB8	4	1.18	9.9	8.00	78.2	
L2251-05	METXB9	5	1.14	9.75	7.93	78.9	
L2251-06	METXC0	6	1.19	9.9	7.82	76.1	
L2251-07	METXC1	7	1.15	9.97	7.59	73.0	
L2251-08	METXC1D	8	1.15	9.97	7.59	73.0	
L2251-09	METXC1S	9	1.15	9.97	7.59	73.0	
L2251-10	METXC3	10	1.11	9.88	8.24	81.3	
L2251-11	METXC4	11	1.14	9.95	8.14	79.5	
L2251-12	METXC5	12	1.13	9.53	7.81	79.5	
L2251-13	METXC6	13	1.18	9.62	7.71	77.4	
L2251-14	METXC7	14	1.16	9.74	7.93	78.9	
L2251-15	METXC8	15	1.18	9.73	7.93	78.9	
L2251-16	METXC9	16	1.16	9.75	8.2	82.0	
L2251-17	METXD0	17	1.18	9.52	7.66	77.7	
L2251-18	METXD1	18	1.18	9.83	7.85	77.1	
L2251-19	METXD2	19	1.15	9.57	7.47	75.1	
L2251-20	METXD3	20	1.17	9.51	7.77	79.1	
L2251-21	METXD4	21	1.12	9.74	7.88	78.4	
L2251-22	METXD5	22	1.11	9.96	7.72	74.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$