


LB108751

7470A

INSTRUMENT ID: CVA



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-59	0.000	-59				
0.2	0.200	0.175	-0.025	701	0.0 %	701				
2.5	2.500	2.477	-0.023	9558	0.0 %	9558				
5.0	5.000	5.048	0.048	19453	0.0 %	19453				
7.5	7.500	7.688	0.188	29608	0.0 %	29608				
10.0	10.000	9.835	-0.165	37870	0.0 %	37870				

LB108751 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Apr 2020 14:33:31

Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-59	-	0.00007470A	PPB	16 Apr 2020 15:15:10	S	Std
0.2 - 1	50.2	701	-	0.20007470A	PPB	16 Apr 2020 15:17:17	S	Std
2.5 - 1	52.5	9558	-	2.50007470A	PPB	16 Apr 2020 15:19:26	S	Std
5.0 - 1	55	19453	-	5.00007470A	PPB	16 Apr 2020 15:21:33	S	Std
7.5 - 1	57.5	29608	-	7.50007470A	PPB	16 Apr 2020 15:23:39	S	Std
10.0 - 1	57.5	37870	-	10.00007470A	PPB	16 Apr 2020 15:25:46	S	Std
ICV01 - 1	ICV01	15032	3.8994	-7470A	PPB	16 Apr 2020 15:28:23	U	SMPL
ICB01 - 1	ICB01	-110	-0.0358	-7470A	PPB	16 Apr 2020 15:30:29	U	SMPL
CCV32 - 1	CCV32	19109	4.9590	-7470A	PPB	16 Apr 2020 15:32:34	U	SMPL
CCB32 - 1	CCB32	-104	-0.0342	-7470A	PPB	16 Apr 2020 15:34:40	U	SMPL
CRA - 1	CRA	715	0.1786	-7470A	PPB	16 Apr 2020 15:36:45	U	SMPL
HighStd - 1	HighStd	37438	9.7225	-7470A	PPB	16 Apr 2020 15:38:51	U	SMPL
ChkStd - 1	ChkStd	26092	6.7738	-7470A	PPB	16 Apr 2020 15:40:57	U	SMPL
PB128284BL - 1	PBW	-122	-0.0389	-7470A	PPB	16 Apr 2020 15:43:05	U	SMPL
PB128284BS - 1	LCSW	14552	3.7747	-7470A	PPB	16 Apr 2020 15:45:13	U	SMPL
L2284-04 - 1	TP-11-SA	-77	-0.0272	-7470A	PPB	16 Apr 2020 15:47:20	U	SMPL
L2284-04DUP - 1	TP-11-SADUP	28	0.0001	-7470A	PPB	16 Apr 2020 15:49:27	U	SMPL
L2284-04MS - 1	TP-11-SAMS	14186	3.6796	-7470A	PPB	16 Apr 2020 15:51:35	U	SMPL
L2284-04MSD - 1	TP-11-SAMSD	14866	3.8563	-7470A	PPB	16 Apr 2020 15:53:43	U	SMPL
L2284-08 - 1	TP-11-DA	119	0.0237	-7470A	PPB	16 Apr 2020 15:55:49	U	SMPL
CCV33 - 1	CCV33	19384	5.0305	-7470A	PPB	16 Apr 2020 15:57:55	U	SMPL
CCB33 - 1	CCB33	-86	-0.0296	-7470A	PPB	16 Apr 2020 16:00:01	U	SMPL
L2284-12 - 1	TP-12-SA	24	-0.0010	-7470A	PPB	16 Apr 2020 16:02:06	U	SMPL
L2284-16 - 1	TP-12-DA	104	0.0198	-7470A	PPB	16 Apr 2020 16:04:12	U	SMPL
L2284-20 - 1	TP-16-B-SA	0	-0.0072	-7470A	PPB	16 Apr 2020 16:06:18	U	SMPL
L2287-05 - 1	GLOUCESTER-SP	10	-0.0046	-7470A	PPB	16 Apr 2020 16:08:25	U	SMPL
L2298-01 - 1	SOIL-CUTTINGS	27	-0.0002	-7470A	PPB	16 Apr 2020 16:10:31	U	SMPL
L2300-01 - 1	WATER-RECLAIM-SYSTEM-SLUDGE-SEPARATOR	236	0.0541	-7470A	PPB	16 Apr 2020 16:12:38	U	SMPL
L2300-02 - 1	WATER-RECLAIM-SYSTEM-SEDIMENT-TANK	107	0.0206	-7470A	PPB	16 Apr 2020 16:14:46	U	SMPL
L2303-01 - 1	RO-11-07	23	-0.0012	-7470A	PPB	16 Apr 2020 16:16:54	U	SMPL
L2303-03 - 1	RO-01-12-11931	40	0.0032	-7470A	PPB	16 Apr 2020 16:19:01	U	SMPL
L2304-02 - 1	WC-B1	48	0.0053	-7470A	PPB	16 Apr 2020 16:21:07	U	SMPL
CCV34 - 1	CCV34	18793	4.8769	-7470A	PPB	16 Apr 2020 16:23:13	U	SMPL
CCB34 - 1	CCB34	-71	-0.0257	-7470A	PPB	16 Apr 2020 16:25:19	U	SMPL
L2304-06 - 1	WC-B2	28	0.0001	-7470A	PPB	16 Apr 2020 16:27:26	U	SMPL
L2304-10 - 1	WC-F1	49	0.0055	-7470A	PPB	16 Apr 2020 16:29:32	U	SMPL
L2304-14 - 1	WC-F2	22	-0.0015	-7470A	PPB	16 Apr 2020 16:31:38	U	SMPL
L2307-03 - 1	TP-10	19	-0.0023	-7470A	PPB	16 Apr 2020 16:33:45	U	SMPL
PB128286BL - 1	PBW	10	-0.0046	-7470A	PPB	16 Apr 2020 16:35:51	U	SMPL
PB128286BS - 1	LCSW	14565	3.7781	-7470A	PPB	16 Apr 2020 16:37:58	U	SMPL
L2271-14 - 1	FB-2020-04-13	-75	-0.0267	-7470A	PPB	16 Apr 2020 16:40:05	U	SMPL
L2279-07 - 1	OILY-WATER	31	0.0009	-7470A	PPB	16 Apr 2020 16:42:13	U	SMPL
L2283-08 - 1	FB-2020-04-14	7	-0.0054	-7470A	PPB	16 Apr 2020 16:44:21	U	SMPL
L2298-02 - 1	GROUNDWATER	41	0.0035	-7470A	PPB	16 Apr 2020 16:46:27	U	SMPL
CCV35 - 1	CCV35	19088	4.9535	-7470A	PPB	16 Apr 2020 16:48:33	U	SMPL
CCB35 - 1	CCB35	-88	-0.0301	-7470A	PPB	16 Apr 2020 16:50:39	U	SMPL
L2298-02DUP - 1	GROUNDWATERDUP	25	-0.0007	-7470A	PPB	16 Apr 2020 16:52:45	U	SMPL
L2298-02MS - 1	GROUNDWATERMS	14612	3.7903	-7470A	PPB	16 Apr 2020 16:54:51	U	SMPL
L2298-02MSD - 1	GROUNDWATERMSD	13666	3.5444	-7470A	PPB	16 Apr 2020 16:56:58	U	SMPL
L2299-01 - 1	COMP-40860-861	2159	0.5539	-7470A	PPB	16 Apr 2020 16:59:04	U	SMPL
L2299-04 - 1	40862	207	0.0466	-7470A	PPB	16 Apr 2020 17:01:10	U	SMPL
L2299-05 - 1	COMP-40863-864-865	905	0.2280	-7470A	PPB	16 Apr 2020 17:03:18	U	SMPL
CCV36 - 1	CCV36	18924	4.9109	-7470A	PPB	16 Apr 2020 17:05:25	U	SMPL
CCB36 - 1	CCB36	-48	-0.0197	-7470A	PPB	16 Apr 2020 17:07:33	U	SMPL
L2284-04LX5 - 1		22	-0.0015	-7470A	PPB	16 Apr 2020 17:09:41	U	SMPL
L2284-04A - 1		14598	3.7866	-7470A	PPB	16 Apr 2020 17:11:47	U	SMPL
L2298-02LX5 - 1		-43	-0.0184	-7470A	PPB	16 Apr 2020 17:13:53	U	SMPL
L2298-02A - 1		14688	3.8100	-7470A	PPB	16 Apr 2020 17:15:59	U	SMPL
PB128284TB - 1	PB128249TB	-44	-0.0186	-7470A	PPB	16 Apr 2020 17:18:05	U	SMPL
CCV37 - 1	CCV37	18813	4.8821	-7470A	PPB	16 Apr 2020 17:20:12	U	SMPL
CCV37 - 1	CCV37	-104	-0.0342	-7470A	PPB	16 Apr 2020 17:22:18	U	SMPL