



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 04/18/2020OVEN IN TEMP1: 98 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 04/17/2020 16:00 DATE & TIME OUT1: 04/18/2020 07:15OVEN IN TEMP2: 104 °COVEN OUT TEMP2: 104 °CDATE & TIME IN2: 04/18/2020 08:00 DATE & TIME OUT2: 04/18/2020 09:35Run Number: LB108763BalanceID: WC SC-3OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB108763BL	LB108763BL	95.3257	95.3257	100	95.3527	95.3527	95.3527	0.0270	270
2	L2306-03	CC-4-15-2020	90.0594	90.0594	100	90.1210	90.1210	90.1210	0.0616	616
3	L2306-03DUP	CC-4-15-2020DUP	89.9999	89.9999	100	90.0636	90.0636	90.0636	0.0637	637

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

78108763

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS L2306 WorkList ID : 137979 Department : Wet-Chemistry Date : 04-17-2020 08:04:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2020	Solid	L2306-03	ASTM TS	Cool 4 deg C	PSEG01	H42	CC-4-15-2020	04/15/2020	SM2540 B

Date/Time 04-17-2020 14:00
Received by: JD
Relinquished by: JD

Date/Time 04-17-2020 16:15
Received by: JD
Relinquished by: JD