



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/18/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 01:01
Out Date: 04/18/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108773

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2287-01	GLOUCESTER-SP	1	1.12	9.6	8.31	84.8	
L2310-01	CLARKSVILLE-TRANSFORMER	2	1.00	2.00	2.00	100.0	OILY-STONE SAMPLE,100% SOLIDS
L2311-01	72-11986	3	1.17	9.58	4.97	45.2	
L2312-01	40693-40720-40531	4	1.00	2.00	2.00	100.0	OILY-DEBRIS,100% SOLIDS
L2327-01	TP-1-S1	5	1.18	9.98	9.14	90.5	
L2327-02	TP-1-S2	6	1.17	9.84	9.32	94.0	
L2327-04	TP-1-D1	7	1.16	9.88	9.31	93.5	
L2327-05	TP-1-D2	8	1.11	9.67	9.38	96.6	
L2327-07	TP-2-S1	9	1.18	9.88	9.48	95.4	
L2327-08	TP-2-S2	10	1.16	9.74	9.38	95.8	
L2327-10	TP-4-S1	11	1.14	9.9	8.8	87.4	
L2327-11	TP-4-S2	12	1.13	9.6	8.56	87.7	
L2327-13	TP-4-D1	13	1.18	9.72	8.43	84.9	
L2327-14	TP-4-D2	14	1.16	9.98	8.69	85.4	
L2327-16	TP-4-S1	15	1.19	9.96	9.21	91.4	
L2327-17	TP-4-S2	16	1.16	9.95	9.28	92.4	
L2327-19	TP-3-S1	17	1.15	9.82	9.41	95.3	
L2327-20	TP-7-S1	18	1.11	9.93	9.5	95.1	
L2327-22	TP-6-S1	19	1.15	9.62	8.96	92.2	
L2327-23	TP-6-S2	20	1.14	9.82	8.99	90.4	
L2329-05	SVOC-GPC-BLANK	21	1.00	2.00	2.00	100.0	
L2329-06	PEST-GPC-BLANK	22	1.00	2.00	2.00	100.0	
L2329-07	PEST-GPC-BLANK-SPIKE	23	1.00	2.00	2.00	100.0	
L2329-08	PCB-GPC-BLANK-	24	1.00	2.00	2.00	100.0	
L2329-09	PCP-GPC-BLANK-SPIKE	25	1.00	2.00	2.00	100.0	
L2329-10	SVOC-GPC2-BLANK	26	1.00	2.00	2.00	100.0	
L2329-11	PEST-GCP2-BLANK	27	1.00	2.00	2.00	100.0	
L2329-12	PEST-GCP2-BLANK-SPIKE	28	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/18/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 01:01
Out Date: 04/18/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108773

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2329-13	PCB-GCP2-BLANK	29	1.00	2.00	2.00	100.0	
L2329-14	PCB-GCP2-BLANK-SPIKE	30	1.00	2.00	2.00	100.0	
L2330-01	WB-9-1	31	1.15	9.52	8.63	89.4	
L2330-02	WB-9-2	32	1.13	9.75	8.89	90.0	
L2330-03	WB-10-1	33	1.11	9.69	8.64	87.8	
L2330-04	WB-10-2	34	1.13	9.91	8.97	89.3	
L2331-01	TP-22- (1.5-2.5)	35	1.13	9.79	8.48	84.9	
L2331-02	TP-22A	36	1.15	9.95	8.89	88.0	
L2331-03	TP-22B	37	1.18	9.68	8.53	86.5	
L2332-01	TP-8-1	38	1.17	9.82	9.17	92.5	
L2332-02	TP-8-2	39	1.18	9.68	8.22	82.8	
L2332-04	TP-4-1	40	1.16	9.58	8.42	86.2	
L2332-05	TP-4-2	41	1.16	9.97	8.8	86.7	
L2337-13	BP-VPB180-378-380	42	1.17	8.9	1.54	4.8	sludge sample
L2339-01	BIN-4	43	1.13	9.57	9.00	93.2	
L2339-02	BIN-4	44	1.14	9.83	9.06	91.1	
L2342-01	TP-18	45	1.12	6.1	5.04	78.7	
L2342-02	TP-18-7.5-8.0	46	1.19	9.77	7.98	79.1	
L2342-03	TP-18-5.5-6.0	47	1.11	9.7	7.97	79.9	
L2342-04	TP-18-3.5-4.0	48	1.12	9.92	8.25	81.0	
L2342-05	TP-18-COMP	49	1.19	9.94	7.94	77.1	
L2342-08	TP-17-16	50	1.16	5.88	5.18	85.2	
L2342-09	TP-17-16-2.5-3.0	51	1.14	9.98	8.68	85.3	
L2342-10	TP-17-16-2.0-2.5	52	1.18	9.61	7.9	79.7	
L2342-11	TP-17-16-1.5-2.0	53	1.19	9.93	8.54	84.1	
L2342-12	TP-17-16-COMP	54	1.17	9.92	8.33	81.8	
L2342-15	TP-15	55	1.11	6.49	5.45	80.7	
L2342-16	TP-15-1.5-2.0-1	56	1.17	9.8	8.22	81.7	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/18/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 01:01
Out Date: 04/18/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108773

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2342-17	TP-15-1.5-2.0-2	57	1.13	9.58	7.9	80.1	
L2342-18	TP-15-1.5-2.0-3	58	1.19	9.84	8.29	82.1	
L2342-19	TP-15-COMP	59	1.14	9.97	8.46	82.9	
L2343-01	SB-3	60	1.16	9.68	7.27	71.7	
L2343-02	SB-1B	61	1.11	9.8	7.12	69.2	
L2343-03	SB-5	62	1.13	9.89	7.43	71.9	
L2343-04	SB-4	63	1.13	9.59	6.93	68.6	
L2343-05	SB-2B	64	1.16	9.77	7.77	76.8	
L2343-06	SB-6	65	1.14	9.95	7.49	72.1	
L2344-01	TP-2-1	66	1.13	9.82	8.9	89.4	
L2344-02	TP-2-2	67	1.17	9.81	8.88	89.2	
L2345-01	T-1	68	1.1	9.86	8.6	85.6	
L2346-01	S-2	85	1.15	9.94	8.76	86.6	
L2346-03	S-3	86	1.13	9.58	8.47	86.9	
L2346-05	S-4	87	1.17	9.66	8.58	87.3	
L2346-07	S-5	88	1.14	9.54	8.89	92.3	
L2346-09	S-6	89	1.18	9.98	8.98	88.6	
L2347-01	TP-4	69	1.12	4.85	3.71	69.4	
L2347-02	TP-4-7.5-8.0	70	1.17	9.64	7.00	68.8	
L2347-03	TP-4-5.5-6.0	71	1.15	9.96	6.91	65.4	
L2347-04	TP-4-3.5-4.0	72	1.1	9.63	6.92	68.2	
L2347-05	TP-4-COMP	73	1.15	9.97	7.28	69.5	
L2348-01	WASTE-1	74	1.15	9.95	8.75	86.4	
L2348-02	VOC	75	1.14	9.93	9.00	89.4	
L2348-03	1	76	1.13	9.77	8.34	83.4	
L2348-04	2	77	1.13	9.81	8.58	85.8	
L2348-05	3	78	1.19	9.97	9.26	91.9	
L2348-06	4	79	1.14	9.88	8.6	85.4	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/18/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 04/17/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 01:01
Out Date: 04/18/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108773

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2348-07	5	80	1.1	9.76	8.78	88.7	
L2350-01	TP-5-S1	81	1.14	9.9	8.94	89.0	
L2350-02	TP-5-S2	82	1.16	9.96	9.42	93.9	
L2350-04	TP-5-D1	83	1.13	9.98	9.44	93.9	
L2350-05	TP-5-D2	84	1.1	9.77	9.09	92.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$