

LB108789


7471B

INSTRUMENT ID: CY1

Linear

 μ Abs.:

48572



A= 0.0000e+000

B= 2.0541e-004

C= -1.7055e-002

Rho= 0.9998962

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	51	0.000	51				
0.2	0.200	0.189	-0.011	1003	0.0 %	1003				
2.5	2.500	2.547	0.047	12482	0.0 %	12482				
5.0	5.000	4.925	-0.075	24060	0.0 %	24060				
7.5	7.500	7.586	0.086	37014	0.0 %	37014				
10.0	10.000	9.960	-0.040	48572	0.0 %	48572				

LB108789 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Apr 2020 10:11:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.00007471B	PPB		20 Apr 2020 10:18:57	S	Std
0.2 - 1	50.2	1003	-	0.20007471B	PPB		20 Apr 2020 10:23:47	S	Std
2.5 - 1	52.5	12482	-	2.50007471B	PPB		20 Apr 2020 10:26:04	S	Std
5.0 - 1	55	24060	-	5.00007471B	PPB		20 Apr 2020 10:28:21	S	Std
7.5 - 1	57.5	37014	-	7.50007471B	PPB		20 Apr 2020 10:33:18	S	Std
10.0 - 1	510	48572	-	10.00007471B	PPB		20 Apr 2020 10:38:10	S	Std
ICV04 - 1	ICV04	20557	4.2055	-7471B	PPB		20 Apr 2020 10:40:54	U	SMPL
ICB04 - 1	ICB04	-15	-0.0201	-7471B	PPB		20 Apr 2020 10:43:09	U	SMPL
CCV51 - 1	CCV51	25003	5.1187	-7471B	PPB		20 Apr 2020 10:47:55	U	SMPL
CCB51 - 1	CCB51	-40	-0.0253	-7471B	PPB		20 Apr 2020 10:50:10	U	SMPL
CRA - 1	CRA	1011	0.1906	-7471B	PPB		20 Apr 2020 10:52:25	U	SMPL
HighStd - 1	HighStd	48584	9.9624	-7471B	PPB		20 Apr 2020 10:54:40	U	SMPL
ChkStd - 1	ChkStd	36875	7.5573	-7471B	PPB		20 Apr 2020 10:56:56	U	SMPL
PB128332BL - 1	PBS	-145	-0.0468	-7471B	PPB		20 Apr 2020 10:59:13	U	SMPL
PB128332BS - 1	LCSS	21651	4.4302	-7471B	PPB		20 Apr 2020 11:01:30	U	SMPL
L2312-01 - 1	40693-40720-40531	752	0.1374	-7471B	PPB		20 Apr 2020 11:03:47	U	SMPL
L2330-01 - 1	WB-9-1	34833	7.1379	-7471B	PPB		20 Apr 2020 11:06:03	U	SMPL
L2330-02 - 1	WB-9-2	53393	10.9502	-7471B	PPB		20 Apr 2020 11:08:20	U	SMPL
L2330-03 - 1	WB-10-1	86517	17.7541	-7471B	PPB		20 Apr 2020 11:10:39	U	SMPL
L2330-04 - 1	WB-10-2	10109	2.0594	-7471B	PPB		20 Apr 2020 11:12:57	U	SMPL
CCV52 - 1	CCV52	23520	4.8141	-7471B	PPB		20 Apr 2020 11:18:53	U	SMPL
CCB52 - 1	CCB52	-17	-0.0205	-7471B	PPB		20 Apr 2020 11:21:09	U	SMPL
L2331-02 - 1	TP-22A	1131	0.2153	-7471B	PPB		20 Apr 2020 11:23:25	U	SMPL
L2331-02DUP - 1	TP-22ADUP	1197	0.2288	-7471B	PPB		20 Apr 2020 11:25:40	U	SMPL
L2331-02MS - 1	TP-22AMS	21309	4.3599	-7471B	PPB		20 Apr 2020 11:27:55	U	SMPL
L2331-02MSD - 1	TP-22AMSD	21142	4.3256	-7471B	PPB		20 Apr 2020 11:30:11	U	SMPL
L2339-01 - 1	BIN-4	3212	0.6427	-7471B	PPB		20 Apr 2020 11:32:27	U	SMPL
L2342-05 - 1	TP-18-COMP	1738	0.3399	-7471B	PPB		20 Apr 2020 11:34:44	U	SMPL
L2342-12 - 1	TP-17-16-COMP	485	0.0826	-7471B	PPB		20 Apr 2020 11:37:01	U	SMPL
L2342-19 - 1	TP-15-COMP	2221	0.4392	-7471B	PPB		20 Apr 2020 11:39:17	U	SMPL
PB128333BL - 1	PBS	9	-0.0152	-7471B	PPB		20 Apr 2020 11:41:35	U	SMPL
PB128333BS - 1	LCSS	18046	3.6897	-7471B	PPB		20 Apr 2020 11:43:50	U	SMPL
CCV53 - 1	CCV53	24902	5.0980	-7471B	PPB		20 Apr 2020 11:46:05	U	SMPL
CCB53 - 1	CCB53	-69	-0.0312	-7471B	PPB		20 Apr 2020 11:48:21	U	SMPL
L2343-01 - 1	SB-3	1946	0.3827	-7471B	PPB		20 Apr 2020 11:50:37	U	SMPL
L2343-02 - 1	SB-1B	3546	0.7113	-7471B	PPB		20 Apr 2020 11:52:52	U	SMPL
L2343-03 - 1	SB-5	1905	0.3742	-7471B	PPB		20 Apr 2020 11:55:08	U	SMPL
L2343-04 - 1	SB-4	1801	0.3529	-7471B	PPB		20 Apr 2020 11:57:23	U	SMPL
L2345-01 - 1	T-1	2858	0.5700	-7471B	PPB		20 Apr 2020 11:59:39	U	SMPL
L2346-01 - 1	S-2	685	0.1236	-7471B	PPB		20 Apr 2020 12:01:56	U	SMPL
L2346-03 - 1	S-3	1378	0.2660	-7471B	PPB		20 Apr 2020 12:04:12	U	SMPL
L2346-05 - 1	S-4	791	0.1454	-7471B	PPB		20 Apr 2020 12:06:29	U	SMPL
L2346-07 - 1	S-5	854	0.1584	-7471B	PPB		20 Apr 2020 12:08:46	U	SMPL
L2346-09 - 1	S-6	731	0.1331	-7471B	PPB		20 Apr 2020 12:11:01	U	SMPL
CCV54 - 1	CCV54	23832	4.8782	-7471B	PPB		20 Apr 2020 12:15:52	U	SMPL
CCB54 - 1	CCB54	-41	-0.0255	-7471B	PPB		20 Apr 2020 12:18:07	U	SMPL
L2347-05 - 1	TP-4-COMP	583	0.1027	-7471B	PPB		20 Apr 2020 12:25:16	U	SMPL
L2347-05DUP - 1	TP-4-COMPDUP	501	0.0859	-7471B	PPB		20 Apr 2020 12:27:32	U	SMPL
L2347-05MS - 1	TP-4-COMPMS	19881	4.0666	-7471B	PPB		20 Apr 2020 12:29:47	U	SMPL
L2347-05MSD - 1	TP-4-COMPMSD	18695	3.8230	-7471B	PPB		20 Apr 2020 12:32:03	U	SMPL
L2331-02LX5 - 1		142	0.0121	-7471B	PPB		20 Apr 2020 12:34:18	U	SMPL
L2331-02A - 1		20693	4.2334	-7471B	PPB		20 Apr 2020 12:36:35	U	SMPL
L2347-05LX5 - 1		79	-0.0008	-7471B	PPB		20 Apr 2020 12:38:52	U	SMPL
L2347-05A - 1		19300	3.9473	-7471B	PPB		20 Apr 2020 12:41:08	U	SMPL
L2330-02DLX2 - 1		26967	5.5221	-7471B	PPB		20 Apr 2020 12:43:25	U	SMPL
L2330-03DLX5 - 1		17329	3.5424	-7471B	PPB		20 Apr 2020 12:45:42	U	SMPL
CCV55 - 1	CCV55	22578	4.6206	-7471B	PPB		20 Apr 2020 12:47:59	U	SMPL
CCB55 - 1	CCB55	-54	-0.0281	-7471B	PPB		20 Apr 2020 12:50:14	U	SMPL