



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108804

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/20/2020	14:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/20/2020	14:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/20/2020	14:18
4	ICV	1	Water	NA	NA	20.1	6.99	04/20/2020	14:19
5	CCV1	1	Water	NA	NA	20.2	2.02	04/20/2020	14:20
6	L2349-01	1	Water	NA	NA	20.6	9.19	04/20/2020	14:22
7	L2349-01DUP	1	Water	NA	NA	20.7	9.20	04/20/2020	14:23
8	CCV2	1	Water	NA	NA	20.1	12.02	04/20/2020	14:25

WORKLIST(Hardcopy Internal Chain)

108809

WorkList Name : CORROSIONITY L2349 WorkList ID : 138059 Department : Wet-Chemistry Date : 04-20-2020 14:05:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/16/2020	Water	L2349-01- ^B	Corrosivity	Cool 4 deg C	URSC05	I51	MS-1	04/16/2020	9040C

Date/Time 04-20-2020 14:10
 Received by: Ja
 Relinquished by: DP

Date/Time 04-20-2020 15:20
 Received by: DP
 Relinquished by: Jn