



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108805

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/20/2020	14:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/20/2020	14:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/20/2020	14:18
4	ICV	1	Water	NA	NA	20.1	6.99	04/20/2020	14:19
5	CCV1	1	Water	NA	NA	20.2	2.02	04/20/2020	14:30
6	L2354-01	1	Water	NA	NA	20.9	5.22	04/20/2020	14:32
7	L2354-01DUP	1	Water	NA	NA	20.8	5.23	04/20/2020	14:33
8	CCV2	1	Water	NA	NA	20.1	12.02	04/20/2020	14:35

108805

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2351 WorkList ID : 138060 Department : Wet-Chemistry Date : 04-20-2020 14:05:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/10/2020	Water	L2354-01	pH	Cool 4 deg C	STER10	I53	CELL-A -POOL-WATER(DE-IC	04/10/2020	SM 4500-H B

Date/Time 04-20-2020 14:11:00
Received by: JO
Relinquished by: DP

Date/Time 04-20-2020 15:20
Received by: DP
Relinquished by: JO