



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108807

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/20/2020	17:15
2	CAL2	1	Water	NA	NA	20.2	7.02	04/20/2020	17:16
3	CAL3	1	Water	NA	NA	20.2	10.03	04/20/2020	17:18
4	ICV	1	Water	NA	NA	20.02	7.00	04/20/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	04/20/2020	17:45
6	L2355-01	1	Water	NA	NA	18.1	5.96	04/20/2020	17:46
7	L2355-01DUP	1	Water	NA	NA	18.2	5.97	04/20/2020	17:48
8	L2355-02	1	Water	NA	NA	19.1	5.53	04/20/2020	17:50
9	L2355-03	1	Water	NA	NA	19.4	6.15	04/20/2020	18:00
10	CCV2	1	Water	NA	NA	20.1	12.02	04/20/2020	18:02

WORKLIST(Hardcopy Internal Chain)

18108807

WorkList Name : PH L2355 WorkList ID : 138063 Department : Wet-Chemistry Date : 04-20-2020 17:03:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/20/2020	Water	L2355-01 ✓	pH	Cool 4 deg C	PSEG01	I41	COMP-32323-32324-32325	04/20/2020	9040C
04/20/2020	Water	L2355-02 ✓	pH	Cool 4 deg C	PSEG01	I41	32326-32327-32328	04/20/2020	9040C
04/20/2020	Water	L2355-03 ✓	pH	Cool 4 deg C	PSEG01	I41	32329-32330-32331-32332	04/20/2020	9040C

Date/Time 04-20-2020 17:10
 Received by: J
 Relinquished by: J

Date/Time 04-20-2020 18:15
 Received by: J
 Relinquished by: J