


LB108809

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	2	0.000	2				
0.2	0.200	0.192	-0.008	714	0.0 %	714				
2.5	2.500	2.527	0.027	8743	0.0 %	8743				
5.0	5.000	4.950	-0.050	17074	0.0 %	17074				
7.5	7.500	7.622	0.122	26261	0.0 %	26261				
10.0	10.000	9.925	-0.075	34178	0.0 %	34178				

LB108809 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 21 Apr 2020 08:36:04					
Sample ID	Extended ID	μ Abs	Conc.	Std Conc/Method	Units	Date	Type	Type	
0.0 - 1	50	2	-	0.0007470A	PPB	21 Apr 2020 08:49:01	S	Std	
0.2 - 1	50.2	714	-	0.20007470A	PPB	21 Apr 2020 08:51:08	S	Std	
2.5 - 1	52.5	8743	-	2.50007470A	PPB	21 Apr 2020 08:53:16	S	Std	
5.0 - 1	55	17074	-	5.00007470A	PPB	21 Apr 2020 08:55:23	S	Std	
7.5 - 1	57.5	26261	-	7.50007470A	PPB	21 Apr 2020 08:57:31	S	Std	
10.0 - 1	510	34178	-	10.00007470A	PPB	21 Apr 2020 08:59:37	S	Std	
ICV05 - 1	ICV05	12817	3.7119	-7470A	PPB	21 Apr 2020 09:02:29	U	SMPL	
ICB05 - 1	ICB05	-120	-0.0508	-7470A	PPB	21 Apr 2020 09:04:34	U	SMPL	
CCV56 - 1	CCV56	16688	4.8377	-7470A	PPB	21 Apr 2020 09:06:40	U	SMPL	
CCB56 - 1	CCB56	-120	-0.0508	-7470A	PPB	21 Apr 2020 09:08:45	U	SMPL	
CRA - 1	CRA	643	0.1712	-7470A	PPB	21 Apr 2020 09:10:51	U	SMPL	
HighStd - 1	HighStd	33127	9.6189	-7470A	PPB	21 Apr 2020 09:12:57	U	SMPL	
ChkStd - 1	ChkStd	22920	6.6502	-7470A	PPB	21 Apr 2020 09:15:03	U	SMPL	
PB128354BL - 1	PBW	-14	-0.0199	-7470A	PPB	21 Apr 2020 09:19:40	U	SMPL	
PB128354BS - 1	LCSW	11940	3.4568	-7470A	PPB	21 Apr 2020 09:21:47	U	SMPL	
L2311-01 - 1	72-11986	-101	-0.0452	-7470A	PPB	21 Apr 2020 09:23:54	U	SMPL	
L2323-06 - 1	TP-20-COMP	-8	-0.0182	-7470A	PPB	21 Apr 2020 09:26:01	U	SMPL	
L2323-13 - 1	TP-19-COMP	9	-0.0132	-7470A	PPB	21 Apr 2020 09:28:08	U	SMPL	
L2327-03 - 1	TP-1-S	0	-0.0159	-7470A	PPB	21 Apr 2020 09:30:16	U	SMPL	
L2327-06 - 1	TP-1-D	18	-0.0106	-7470A	PPB	21 Apr 2020 09:32:22	U	SMPL	
CCV57 - 1	CCV57	15584	4.5166	-7470A	PPB	21 Apr 2020 09:34:28	U	SMPL	
CCB57 - 1	CCB57	-102	-0.0455	-7470A	PPB	21 Apr 2020 09:36:33	U	SMPL	
L2327-09 - 1	TP-2-S	-6	-0.0176	-7470A	PPB	21 Apr 2020 09:38:39	U	SMPL	
L2327-12 - 1	TP-4-S	2	-0.0153	-7470A	PPB	21 Apr 2020 09:40:45	U	SMPL	
L2327-15 - 1	TP-4-D	16	-0.0112	-7470A	PPB	21 Apr 2020 09:42:51	U	SMPL	
L2327-18 - 1	TP-4-S	14	-0.0118	-7470A	PPB	21 Apr 2020 09:44:57	U	SMPL	
L2327-21 - 1	TP-3-S	15	-0.0115	-7470A	PPB	21 Apr 2020 09:47:04	U	SMPL	
L2327-24 - 1	TP-6-S	20	-0.0100	-7470A	PPB	21 Apr 2020 09:49:11	U	SMPL	
L2331-04 - 1	TP-22A	-2	-0.0164	-7470A	PPB	21 Apr 2020 09:51:18	U	SMPL	
L2332-03 - 1	TP-8	12	-0.0124	-7470A	PPB	21 Apr 2020 09:53:25	U	SMPL	
L2332-06 - 1	TP-4	22	-0.0095	-7470A	PPB	21 Apr 2020 09:55:33	U	SMPL	
L2340-01 - 1	CHASSIS-WASH	5	-0.0144	-7470A	PPB	21 Apr 2020 09:57:40	U	SMPL	
CCV58 - 1	CCV58	15907	4.6106	-7470A	PPB	21 Apr 2020 10:03:23	U	SMPL	
CCB58 - 1	CCB58	-97	-0.0441	-7470A	PPB	21 Apr 2020 10:05:29	U	SMPL	
L2340-02 - 1	BUS-WASH	33	-0.0063	-7470A	PPB	21 Apr 2020 10:07:35	U	SMPL	
L2340-03 - 1	RECLAIM-SYSTEM-SEDIMENT-PIT	69	0.0042	-7470A	PPB	21 Apr 2020 10:09:43	U	SMPL	
L2344-03 - 1	TP-2	-25	-0.0231	-7470A	PPB	21 Apr 2020 10:11:49	U	SMPL	
L2344-03DUP - 1	TP-2DUP	14	-0.0118	-7470A	PPB	21 Apr 2020 10:13:55	U	SMPL	
L2344-03MS - 1	TP-2MS	13523	3.9172	-7470A	PPB	21 Apr 2020 10:19:59	U	SMPL	
L2344-03MSD - 1	TP-2MSD	13629	3.9480	-7470A	PPB	21 Apr 2020 10:22:06	U	SMPL	
PB128356BL - 1	PBW	-53	-0.0313	-7470A	PPB	21 Apr 2020 10:24:13	U	SMPL	
PB128356BS - 1	LCSW	14650	4.2450	-7470A	PPB	21 Apr 2020 10:26:20	U	SMPL	
L2330-05 - 1	WB-9-1	-55	-0.0319	-7470A	PPB	21 Apr 2020 10:28:28	U	SMPL	
L2330-06 - 1	WB-9-2	23	-0.0092	-7470A	PPB	21 Apr 2020 10:30:36	U	SMPL	
CCV59 - 1	CCV59	16840	4.8819	-7470A	PPB	21 Apr 2020 10:35:21	U	SMPL	
CCB59 - 1	CCB59	-118	-0.0502	-7470A	PPB	21 Apr 2020 10:37:27	U	SMPL	
L2330-07 - 1	WB-10-1	1	-0.0156	-7470A	PPB	21 Apr 2020 10:39:33	U	SMPL	
L2330-08 - 1	WB-10-2	28	-0.0077	-7470A	PPB	21 Apr 2020 10:41:40	U	SMPL	
L2339-03 - 1	BIN-4	18	-0.0106	-7470A	PPB	21 Apr 2020 10:43:46	U	SMPL	
L2341-01 - 1	CHASSIS-WASH-1	42	-0.0036	-7470A	PPB	21 Apr 2020 10:45:52	U	SMPL	
L2341-02 - 1	CHASSIS-WASH-2	90	0.0103	-7470A	PPB	21 Apr 2020 10:47:59	U	SMPL	
L2341-03 - 1	BUS-WASH	100	0.0132	-7470A	PPB	21 Apr 2020 10:50:05	U	SMPL	
L2341-04 - 1	WATER-RECLAIM-SEDIMENT-PIT	72	0.0051	-7470A	PPB	21 Apr 2020 10:52:13	U	SMPL	
L2342-06 - 1	TP-18-COMP	5	-0.0144	-7470A	PPB	21 Apr 2020 10:54:20	U	SMPL	
L2342-13 - 1	TP-17-16-COMP	21	-0.0097	-7470A	PPB	21 Apr 2020 10:56:28	U	SMPL	
L2342-20 - 1	TP-15-COMP	24	-0.0089	-7470A	PPB	21 Apr 2020 10:58:34	U	SMPL	
CCV60 - 1	CCV60	16952	4.9145	-7470A	PPB	21 Apr 2020 11:00:41	U	SMPL	
CCB60 - 1	CCB60	-66	-0.0351	-7470A	PPB	21 Apr 2020 11:02:46	U	SMPL	
L2347-06 - 1	TP-4-COMP	12	-0.0124	-7470A	PPB	21 Apr 2020 11:04:52	U	SMPL	
L2350-03 - 1	TP-5-S	23	-0.0092	-7470A	PPB	21 Apr 2020 11:06:59	U	SMPL	
L2350-06 - 1	TP-5-D	40	-0.0042	-7470A	PPB	21 Apr 2020 11:09:05	U	SMPL	
L2350-06DUP - 1	TP-5-DDUP	39	-0.0045	-7470A	PPB	21 Apr 2020 11:11:11	U	SMPL	
L2350-06MS - 1	TP-5-DMS	13970	4.0472	-7470A	PPB	21 Apr 2020 11:13:18	U	SMPL	
L2350-06MSD - 1	TP-5-DMSD	14265	4.1330	-7470A	PPB	21 Apr 2020 11:15:25	U	SMPL	
PB128357BL - 1	PBW	-78	-0.0385	-7470A	PPB	21 Apr 2020 11:17:32	U	SMPL	
PB128357BS - 1	LCSW	15199	4.4047	-7470A	PPB	21 Apr 2020 11:19:40	U	SMPL	
L2279-08 - 1	OILY-WATER	-79	-0.0388	-7470A	PPB	21 Apr 2020 11:21:48	U	SMPL	
L2279-08DUP - 1	OILY-WATERDUP	14	-0.0118	-7470A	PPB	21 Apr 2020 11:23:54	U	SMPL	
CCV61 - 1	CCV61	17172	4.9785	-7470A	PPB	21 Apr 2020 11:26:01	U	SMPL	
CCB61 - 1	CCB61	-92	-0.0426	-7470A	PPB	21 Apr 2020 11:28:08	U	SMPL	
L2279-08MS - 1	OILY-WATERMS	14800	4.2886	-7470A	PPB	21 Apr 2020 11:30:15	U	SMPL	
L2279-08MSD - 1	OILY-WATERMSD	16147	4.6804	-7470A	PPB	21 Apr 2020 11:32:22	U	SMPL	
PB128360BL - 1	PBW	-94	-0.0432	-7470A	PPB	21 Apr 2020 11:34:29	U	SMPL	
PB128360BS - 1	LCSW	13725	3.9760	-7470A	PPB	21 Apr 2020 11:36:36	U	SMPL	
L2330-09 - 1	FB-GW-2020-04-16	-75	-0.0377	-7470A	PPB	21 Apr 2020 11:38:42	U	SMPL	
L2330-10 - 1	FB-SO-2020-04-16	9	-0.0132	-7470A	PPB	21 Apr 2020 11:40:49	U	SMPL	
L2330-12 - 1	PA-TWP-2	1789	0.5045	-7470A	PPB	21 Apr 2020 11:45:24	U	SMPL	
L2330-12DUP - 1	PA-TWP-2DUP	1584	0.4448	-7470A	PPB	21 Apr 2020 11:47:31	U	SMPL	
L2330-13 - 1	L2330-12MS	16657	4.8287	-7470A	PPB	21 Apr 2020 11:49:39	U	SMPL	
L2330-14 - 1	L2330-12MSD	17970	5.2106	-7470A	PPB	21 Apr 2020 11:51:46	U	SMPL	
CCV62 - 1	CCV62	16754	4.8569	-7470A	PPB	21 Apr 2020 11:56:30	U	SMPL	
CCB62 - 1	CCB62	-46	-0.0292	-7470A	PPB	21 Apr 2020 11:58:36	U	SMPL	
L2330-15 - 1	PA-TWP-3	740	0.1994	-7470A	PPB	21 Apr 2020 12:00:43	U	SMPL	
L2330-16 - 1	DUP-2020-04-16	10020	2.8984	-7470A	PPB	21 Apr 2020 12:02:50	U	SMPL	
L2330-17 - 1	PA-TWP-1	-41	-0.0278	-7470A	PPB	21 Apr 2020 12:04:57	U	SMPL	
L2330-18 - 1	WB-TWP-1	5880	1.6943	-7470A	PPB	21 Apr 2020 12:07:04	U	SMPL	
L2330-19 - 1	WB-TWP-2	78	0.0068	-7470A	PPB	21 Apr 2020 12:09:11	U	SMPL	
L2330-20 - 1	WB-TWP-3	40	-0.0042	-7470A	PPB	21 Apr 2020 12:11:18	U	SMPL	
L2330-21 - 1	WB-TWP-4	30	-0.0071	-7470A	PPB	21 Apr 2020 12:13:25	U	SMPL	
L2349-01 - 1	MS-1	452	0.1156	-7470A	PPB	21 Apr 2020 12:15:33	U	SMPL	
PB128354TB - 1	PB128312TB	-94	-0.0432	-7470A	PPB	21 Apr 2020 12:17:41	U	SMPL	
PB128356TB - 1	PB128336TB	7	-0.0138	-7470A	PPB	21 Apr 2020 12:19:47	U	SMPL	
CCV63 - 1	CCV63	16876	4.8924	-7470A	PPB	21 Apr 2020 12:21:54	U	SMPL	

Method: 7470A

Operator: Admin

LB108809 INSTRUMENT ID : CV1

Date of Analysis: 21 Apr 2020 08:36:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB63 - 1	CCB63	-36	-0.0263	-7470A	PPB		21 Apr 2020 12:24:01	U	SMPL
PB128357TB - 1	PB128338TB	22	-0.0095	-7470A	PPB		21 Apr 2020 12:26:08	U	SMPL
L2344-03LX5 - 1		11	-0.0127	-7470A	PPB		21 Apr 2020 12:28:15	U	SMPL
L2350-06LX5 - 1		13678	3.9623	-7470A	PPB		21 Apr 2020 12:30:22	U	SMPL
L2350-06A - 1		-35	-0.0260	-7470A	PPB		21 Apr 2020 12:32:29	U	SMPL
L2279-08LX5 - 1		13819	4.0033	-7470A	PPB		21 Apr 2020 12:34:36	U	SMPL
L2279-08A - 1		-41	-0.0278	-7470A	PPB		21 Apr 2020 12:36:42	U	SMPL
L2330-12LX5 - 1		14843	4.3011	-7470A	PPB		21 Apr 2020 12:38:50	U	SMPL
L2330-12A - 1		499	0.1293	-7470A	PPB		21 Apr 2020 12:40:57	U	SMPL
CCV64 - 1	CCV64	15443	4.4756	-7470A	PPB		21 Apr 2020 12:43:05	U	SMPL
CCB64 - 1	CCB64	16080	4.6609	-7470A	PPB		21 Apr 2020 12:45:12	U	SMPL
		-81	-0.0394	-7470A	PPB		21 Apr 2020 12:47:19	U	SMPL