



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: Daniel

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108815

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/21/2020	08:00
2	L2279-08		1	8	100.00	>212.0	>212.0	04/21/2020	08:20
3	L2279-08DUP		1	9	100.00	>212.0	>212.0	04/21/2020	08:45
4	L2286-01		1	8	100.00	>212.0	>212.0	04/21/2020	09:15
5	L2298-02		1	9	100.00	>212.0	>212.0	04/21/2020	09:35
6	L2299-01		1	9	100.00	>212.0	>212.0	04/21/2020	10:00
7	L2299-04		1	8	100.00	>212.0	>212.0	04/21/2020	10:20
8	L2299-05		1	8	100.00	>212.0	>212.0	04/21/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB108815

WorkList Name : fp041720 WorkList ID : 138003 Department : Wet-Chemistry Date : 04-17-2020 11:26:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2020	Water	L2279-08	Flash Point	Cool 4 deg C	PSEG01	H31	OILY-WATER	04/14/2020	1010A
04/28/2020	Water	L2286-01	Flash Point	Cool 4 deg C	DIST01		55-ST-27	04/15/2020	1010A
04/28/2020	Water	L2298-02	Flash Point	Cool 4 deg C	SEAC01	G41	GROUNDWATER	04/14/2020	1010A
04/21/2020	Water	L2299-01	Flash Point	Cool 4 deg C	PSEG01	H41	COMP-40860-861	04/14/2020	1010A
04/21/2020	Water	L2299-04	Flash Point	Cool 4 deg C	PSEG01	H41	40862	04/14/2020	1010A
04/21/2020	Water	L2299-05	Flash Point	Cool 4 deg C	PSEG01	H41	COMP-40863-864-865	04/14/2020	1010A

Date/Time 04/21/2020 7:30
 Received by: Dye
 Relinquished by: AM

Date/Time 04/21/2020 10:50
 Received by: AM
 Relinquished by: Dye