



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108822

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/21/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	04/21/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	04/21/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.02	04/21/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.02	04/21/2020	17:16
6	L2331-08	1	Solid	20.03	20	20.9	10.11	04/21/2020	17:19
7	L2331-08DUP	1	Solid	20.04	20	20.8	10.12	04/21/2020	17:20
8	L2359-04	1	Solid	20.02	20	21.2	6.64	04/21/2020	17:22
9	L2359-08	1	Solid	20.03	20	21.2	6.09	04/21/2020	17:25
10	L2375-04	1	Solid	20.02	20	20.4	6.52	04/21/2020	17:26
11	CCV2	1	Water	NA	NA	20.2	12.02	04/21/2020	17:30