


LB108823

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-3	0.000	-3				
0.2	0.200	0.164	-0.036	743	0.0 %	743				
2.5	2.500	2.444	-0.056	9289	0.0 %	9289				
5.0	5.000	5.079	0.079	19165	0.0 %	19165				
7.5	7.500	7.926	0.426	29833	0.0 %	29833				
10.0	10.000	9.622	-0.378	36191	0.0 %	36191				

LB108823 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Apr 2020 08:36:25					
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.0000	7470A	PPB	22 Apr 2020 08:44:57	S	Std
0.2 - 1	50.2	743	-	0.2000	7470A	PPB	22 Apr 2020 08:47:05	S	Std
2.5 - 1	52.5	9289	-	2.5000	7470A	PPB	22 Apr 2020 08:49:12	S	Std
5.0 - 1	55	19165	-	5.0000	7470A	PPB	22 Apr 2020 08:51:20	S	Std
7.5 - 1	57.5	29833	-	7.5000	7470A	PPB	22 Apr 2020 08:53:27	S	Std
10.0 - 1	511	36191	-	10.0000	7470A	PPB	22 Apr 2020 08:55:33	S	Std
ICV07 - 1	ICV07	14500	3.8345	-	7470A	PPB	22 Apr 2020 08:58:17	U	SMPL
ICB07 - 1	ICB07	-114	-0.0648	-	7470A	PPB	22 Apr 2020 09:00:22	U	SMPL
CCV70 - 1	CCV70	19467	5.1598	-	7470A	PPB	22 Apr 2020 09:02:28	U	SMPL
CCB70 - 1	CCB70	-130	-0.0691	-	7470A	PPB	22 Apr 2020 09:04:34	U	SMPL
CRA - 1	CRA	747	0.1649	-	7470A	PPB	22 Apr 2020 09:06:40	U	SMPL
HighStd - 1	HighStd	36125	9.6046	-	7470A	PPB	22 Apr 2020 09:08:45	U	SMPL
ChkStd - 1	ChkStd	27235	7.2325	-	7470A	PPB	22 Apr 2020 09:10:51	U	SMPL
PB128394BL - 1	PBW	-169	-0.0795	-	7470A	PPB	22 Apr 2020 09:12:59	U	SMPL
PB128394BS - 1	LCSW	13862	3.6643	-	7470A	PPB	22 Apr 2020 09:15:06	U	SMPL
L2123-14 - 1	TR-05-042020	-92	-0.0589	-	7470A	PPB	22 Apr 2020 09:17:13	U	SMPL
L2123-14DUP - 1	TR-05-042020DUP	-2	-0.0349	-	7470A	PPB	22 Apr 2020 09:19:20	U	SMPL
L2123-14MS - 1	TR-05-042020MS	14514	3.8383	-	7470A	PPB	22 Apr 2020 09:21:28	U	SMPL
L2123-14MSD - 1	TR-05-042020MSD	13961	3.6907	-	7470A	PPB	22 Apr 2020 09:23:35	U	SMPL
L2351-04 - 1	TP-17A	-64	-0.0514	-	7470A	PPB	22 Apr 2020 09:25:41	U	SMPL
CCV71 - 1	CCV71	19866	5.2663	-	7470A	PPB	22 Apr 2020 09:27:46	U	SMPL
CCB71 - 1	CCB71	-121	-0.0667	-	7470A	PPB	22 Apr 2020 09:29:52	U	SMPL
L2351-08 - 1	TP-20A	6	-0.0328	-	7470A	PPB	22 Apr 2020 09:31:59	U	SMPL
L2351-12 - 1	TP-21A	19	-0.0293	-	7470A	PPB	22 Apr 2020 09:34:04	U	SMPL
L2353-04 - 1	VNJ-240	31	-0.0261	-	7470A	PPB	22 Apr 2020 09:36:10	U	SMPL
L2356-06 - 1	TP-13-12-COMP	10	-0.0317	-	7470A	PPB	22 Apr 2020 09:38:16	U	SMPL
L2357-04 - 1	TP-6	28	-0.0269	-	7470A	PPB	22 Apr 2020 09:40:22	U	SMPL
PB128394TB - 1	PB128374TB	25	-0.0277	-	7470A	PPB	22 Apr 2020 09:42:29	U	SMPL
L2123-14LX5 - 1		5	-0.0330	-	7470A	PPB	22 Apr 2020 09:44:36	U	SMPL
L2123-14A - 1		14606	3.8628	-	7470A	PPB	22 Apr 2020 09:46:44	U	SMPL
CCV72 - 1	CCV72	19981	5.2970	-	7470A	PPB	22 Apr 2020 09:48:52	U	SMPL
CCB72 - 1	CCB72	-121	-0.0667	-	7470A	PPB	22 Apr 2020 09:50:58	U	SMPL