



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108838

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/22/2020	14:40
2	CAL2	1	Water	NA	NA	20.2	7.00	04/22/2020	14:41
3	CAL3	1	Water	NA	NA	20.2	10.03	04/22/2020	14:42
4	ICV	1	Water	NA	NA	20.2	7.01	04/22/2020	14:43
5	CCV1	1	Water	NA	NA	20.2	2.02	04/22/2020	15:10
6	L2384-01	1	Water	NA	NA	17.7	5.09	04/22/2020	15:11
7	L2384-01DUP	1	Water	NA	NA	17.8	5.10	04/22/2020	15:12
8	CCV2	1	Water	NA	NA	20.1	12.02	04/22/2020	15:15

WORKLIST(Hardcopy Internal Chain)

1088358

Date : 04-22-2020 14:28:38

Department : Wet-Chemistry

WorkList ID : 138138

WorkList Name : PH L2384

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/21/2020	Water	L2384-01	pH	Cool 4 deg C	STER11	J11	POOL-WATER	04/21/2020	SM 4500-H B

Date/Time 04.22.2020 14:30
 Received by: Jn
 Relinquished by: Jn

Date/Time 04.22.2020 15:35
 Received by: Jn
 Relinquished by: Jn