



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108839

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/22/2020	14:40
2	CAL2	1	Water	NA	NA	20.2	7.00	04/22/2020	14:41
3	CAL3	1	Water	NA	NA	20.2	10.03	04/22/2020	14:42
4	ICV	1	Water	NA	NA	20.1	7.01	04/22/2020	14:43
5	CCV1	1	Water	NA	NA	20.1	2.02	04/22/2020	17:15
6	L2117-10	1	Solid	20.03	20	21.1	8.88	04/22/2020	17:16
7	L2120-05	1	Solid	20.02	20	21.1	8.17	04/22/2020	17:19
8	L2324-04	1	Solid	20.03	20	21.2	8.95	04/22/2020	17:20
9	L2378-04	1	Solid	20.04	20	21.0	8.03	04/22/2020	17:22
10	L2378-08	1	Solid	20.04	20	21.5	9.09	04/22/2020	17:25
11	L2378-12	1	Solid	20.02	20	21.2	8.70	04/22/2020	17:30
12	L2397-06	1	Solid	20.01	20	20.5	7.30	04/22/2020	17:32
13	L2397-06DUP	1	Solid	20.03	20	20.6	7.31	04/22/2020	17:33
14	CCV2	1	Water	NA	NA	20.2	12.03	04/22/2020	17:35

WORKLIST(Hardcopy Internal Chain)

✓ 108839

WorkList Name : CORROSIONITY L2397

WorkList ID : 138135

Department : Wet-Chemistry

Date : 04-22-2020 14:25:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/22/2020	Solid	L2117-10	Corrosivity	Cool 4 deg C	PSEG05		OK-01-042220	04/22/2020	9045D
04/22/2020	Solid	L2120-05	Corrosivity	Cool 4 deg C	PSEG05	J12	OR-02-042220	04/22/2020	9045D
04/16/2020	Solid	L2324-04	Corrosivity	Cool 4 deg C	FIRS02		WC-1	04/16/2020	9045D
04/21/2020	Solid	L2378-04	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-6-S	04/21/2020	9045D
04/21/2020	Solid	L2378-08	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-3-S	04/21/2020	9045D
04/21/2020	Solid	L2378-12	Corrosivity	Cool 4 deg C	PSEG01	I32	TP-2-S	04/21/2020	9045D
04/20/2020	Solid	L2397-06	Corrosivity	Cool 4 deg C	PSEG01	J12	TP-1-COMP	04/20/2020	9045D

Date/Time 04-22-2020 16:00

Received by: Jn

Relinquished by: Jn

Date/Time 04-22-2020 18:15

Received by: Jn

Relinquished by: Jn