

LB 108843


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40274



A= 0.0000e+000

B= 2.4745e-004

C= 4.5428e-002

Rho= 0.9995739

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	-47	0.000	-47				
0.2	0.200	0.217	0.017	692	0.0 %	692				
2.5	2.500	2.559	0.059	10159	0.0 %	10159				
5.0	5.000	4.770	-0.230	19095	0.0 %	19095				
7.5	7.500	7.609	0.109	30566	0.0 %	30566				
10.0	10.000	10.011	0.011	40274	0.0 %	40274				

LB108843 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Apr 2020 08:52:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-47	-	0.0000	7470A	PPB	23 Apr 2020 08:55:13	S	Std
0.2 - 1	50.2	692	-	0.2000	7470A	PPB	23 Apr 2020 08:57:21	S	Std
2.5 - 1	52.5	10159	-	2.5000	7470A	PPB	23 Apr 2020 08:59:29	S	Std
5.0 - 1	55	19095	-	5.0000	7470A	PPB	23 Apr 2020 09:01:36	S	Std
7.5 - 1	57.5	30566	-	7.5000	7470A	PPB	23 Apr 2020 09:03:43	S	Std
10.0 - 1	510	40274	-	10.0000	7470A	PPB	23 Apr 2020 09:05:50	S	Std
ICV10 - 1	ICV10	14793	3.7059	-	7470A	PPB	23 Apr 2020 09:08:46	U	SMPL
ICB10 - 1	ICB10	-146	0.0093	-	7470A	PPB	23 Apr 2020 09:10:52	U	SMPL
CCV79 - 1	CCV79	18911	4.7249	-	7470A	PPB	23 Apr 2020 09:12:58	U	SMPL
CCB79 - 1	CCB79	-126	0.0142	-	7470A	PPB	23 Apr 2020 09:15:03	U	SMPL
CRA - 1	CRA	736	0.2275	-	7470A	PPB	23 Apr 2020 09:17:09	U	SMPL
HighStd - 1	HighStd	39408	9.7968	-	7470A	PPB	23 Apr 2020 09:19:14	U	SMPL
ChkStd - 1	ChkStd	25961	6.4694	-	7470A	PPB	23 Apr 2020 09:21:20	U	SMPL
PB128421BL - 1	PBW	-195	-0.0028	-	7470A	PPB	23 Apr 2020 09:23:28	U	SMPL
PB128421BS - 1	LCSW	14766	3.6992	-	7470A	PPB	23 Apr 2020 09:25:35	U	SMPL
L2331-08 - 1	TP-19A	-106	0.0192	-	7470A	PPB	23 Apr 2020 09:27:42	U	SMPL
L2331-08DUP - 1	TP-19ADUP	-7	0.0437	-	7470A	PPB	23 Apr 2020 09:29:50	U	SMPL
L2331-08MS - 1	TP-19AMS	15219	3.8113	-	7470A	PPB	23 Apr 2020 09:31:58	U	SMPL
L2331-08MSD - 1	TP-19AMSD	13186	3.3082	-	7470A	PPB	23 Apr 2020 09:34:05	U	SMPL
L2359-04 - 1	TP-1B-S	-51	0.0328	-	7470A	PPB	23 Apr 2020 09:36:12	U	SMPL
CCV80 - 1	CCV80	19334	4.8295	-	7470A	PPB	23 Apr 2020 09:38:18	U	SMPL
CCB80 - 1	CCB80	-115	0.0170	-	7470A	PPB	23 Apr 2020 09:40:25	U	SMPL
L2359-08 - 1	TP-1B-D	27	0.0521	-	7470A	PPB	23 Apr 2020 09:42:31	U	SMPL
L2375-04 - 1	TP-16A	31	0.0531	-	7470A	PPB	23 Apr 2020 09:44:36	U	SMPL
PB128421TB - 1	PB128386TB	8	0.0474	-	7470A	PPB	23 Apr 2020 09:46:42	U	SMPL
L2331-08LX5 - 1		21	0.0506	-	7470A	PPB	23 Apr 2020 09:48:48	U	SMPL
L2331-08A - 1		14818	3.7121	-	7470A	PPB	23 Apr 2020 09:50:55	U	SMPL
CCV81 - 1	CCV81	19044	4.7578	-	7470A	PPB	23 Apr 2020 09:53:02	U	SMPL
CCB81 - 1	CCB81	-105	0.0194	-	7470A	PPB	23 Apr 2020 09:55:09	U	SMPL