


LB108849

245.1

INSTRUMENT ID: CY1



A= 0.0000e+000
 B= 2.6951e-004
 C= -1.4032e-001
 Rho= 0.9994724
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	586	0.000	586				
0.05	0.050	0.046	-0.004	690	0.0 %	690				
0.2	0.200	0.161	-0.039	1117	0.0 %	1117				
2.5	2.500	2.447	-0.053	9600	0.0 %	9600				
5.0	5.000	5.029	0.029	19180	0.0 %	19180				
7.5	7.500	7.743	0.243	29251	0.0 %	29251				
10.0	10.000	9.807	-0.193	36910	0.0 %	36910				

Method: 245.1

Operator: Admin

LB108849 INSTRUMENT ID : CV1

Date of Analysis: 23 Apr 2020 13:17:07

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	586	-	0.0000245.1	PPB	23 Apr 2020 13:25:11	S	Std
0.05 - 1	50.05	690	-	0.0500245.1	PPB	23 Apr 2020 13:27:28	S	Std
0.2 - 1	50.2	1117	-	0.2000245.1	PPB	23 Apr 2020 13:29:45	S	Std
2.5 - 1	52.5	9600	-	2.5000245.1	PPB	23 Apr 2020 13:32:02	S	Std
5.0 - 1	55	19180	-	5.0000245.1	PPB	23 Apr 2020 13:34:18	S	Std
7.5 - 1	57.5	29251	-	7.5000245.1	PPB	23 Apr 2020 13:36:34	S	Std
10.0 - 1	57.5	36910	-	10.0000245.1	PPB	23 Apr 2020 13:49:12	S	Std
ICV12 - 1	ICV12	15210	3.9589	-245.1	PPB	23 Apr 2020 13:55:11	U	SMPL
ICB12 - 1	ICB12	352	-0.0455	-245.1	PPB	23 Apr 2020 13:57:26	U	SMPL
CCV86 - 1	CCV86	19217	5.0388	-245.1	PPB	23 Apr 2020 13:59:40	U	SMPL
CCB86 - 1	CCB86	388	-0.0357	-245.1	PPB	23 Apr 2020 14:01:55	U	SMPL
CRA - 1	CRA	704	0.0494	-245.1	PPB	23 Apr 2020 14:04:10	U	SMPL
HighStd - 1	HighStd	36829	9.7854	-245.1	PPB	23 Apr 2020 14:06:26	U	SMPL
ChkStd - 1	ChkStd	26590	7.0259	-245.1	PPB	23 Apr 2020 14:08:40	U	SMPL
PB128441BL - 1	PBW	488	-0.0088	-245.1	PPB	23 Apr 2020 14:14:01	U	SMPL
PB128441BS - 1	LCSW	13825	3.5856	-245.1	PPB	23 Apr 2020 14:18:51	U	SMPL
L2182-10 - 1	LOD-MDL-WATER-04-QT2-2020	729	0.0562	-245.1	PPB	23 Apr 2020 14:23:40	U	SMPL
L2182-11 - 1	LOQ-WATER-05-QT2-2020	1138	0.1664	-245.1	PPB	23 Apr 2020 14:25:56	U	SMPL
L2286-01 - 1	55-ST-27	65	-0.1228	-245.1	PPB	23 Apr 2020 14:30:49	U	SMPL
L2286-01DUP - 1	55-ST-27DUP	47	-0.1277	-245.1	PPB	23 Apr 2020 14:33:06	U	SMPL
L2286-01MS - 1	55-ST-27MS	16545	4.3187	-245.1	PPB	23 Apr 2020 14:35:21	U	SMPL
CCV87 - 1	CCV87	19200	5.0342	-245.1	PPB	23 Apr 2020 14:37:36	U	SMPL
CCB87 - 1	CCB87	358	-0.0438	-245.1	PPB	23 Apr 2020 14:39:52	U	SMPL
L2286-01MSD - 1	55-ST-27MSD	16340	4.2634	-245.1	PPB	23 Apr 2020 14:42:07	U	SMPL
L2286-01LX5 - 1		388	-0.0357	-245.1	PPB	23 Apr 2020 14:47:10	U	SMPL
L2286-01A - 1		16356	4.2677	-245.1	PPB	23 Apr 2020 14:49:25	U	SMPL
CCV88 - 1	CCV88	19250	5.0477	-245.1	PPB	23 Apr 2020 14:51:40	U	SMPL
CCB88 - 1	CCB88	353	-0.0452	-245.1	PPB	23 Apr 2020 14:53:56	U	SMPL