



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/24/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 04/23/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 04/24/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108856

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1760-01	ZERO-OUT-INVOICE	1	1.00	2.00	2.00	100.0	
L1993-01	xyz	2	1.00	2.00	2.00	100.0	
L2399-01	SP1-A	3	1.13	9.55	9.04	93.9	
L2399-02	SP1-A-EPH-2	4	1.13	9.55	9.04	93.9	
L2399-06	SP1-B	5	1.14	9.62	9.06	93.4	
L2399-07	SP1-B-EPH-2	6	1.14	9.62	9.06	93.4	
L2401-01	DRAINAGE-SEWALE	7	1.13	9.87	8.94	89.4	
L2401-03	DRAINAGE-SEWALE	8	1.18	9.51	8.83	91.8	
L2401-05	CONCRETE-DEMO-1	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE,100% SOLIDS
L2401-06	CONCRETE-DEMO-2	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE,100% SOLIDS
L2402-01	01-A-01-B-01-C	11	1.00	2.00	2.00	100.0	CAULKING sample,100% solids
L2402-02	02-A-02-B-02-C	12	1.00	2.00	2.00	100.0	CAULKING sample,100% solids
L2404-01	WTP-RESIN-01	13	1.17	9.62	6.33	61.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$