

L3108861


7470A

INSTRUMENT ID : PY2

Linear

μ Abs. :

37699



A= 0.0000e+000

B= 2.6188e-004

C= -1.3969e-002

Rho= 0.9995606

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.030	0.030	168	0.000	168				
0.2	0.200	0.192	-0.008	786	0.0 %	786				
2.5	2.500	2.407	-0.093	9244	0.0 %	9244				
5.0	5.000	5.007	0.007	19174	0.0 %	19174				
7.5	7.500	7.705	0.205	29476	0.0 %	29476				
10.0	10.000	9.859	-0.141	37699	0.0 %	37699				

LB108861 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2020 09:11:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	168	-	0.0000	7470A	PPB	24 Apr 2020 09:13:37	S	Std
0.2 - 1	50.2	786	-	0.2000	7470A	PPB	24 Apr 2020 09:15:45	S	Std
2.5 - 1	52.5	9244	-	2.5000	7470A	PPB	24 Apr 2020 09:17:52	S	Std
5.0 - 1	55	19174	-	5.0000	7470A	PPB	24 Apr 2020 09:19:59	S	Std
7.5 - 1	57.5	29476	-	7.5000	7470A	PPB	24 Apr 2020 09:22:06	S	Std
10.0 - 1	57.5	37699	-	10.0000	7470A	PPB	24 Apr 2020 09:24:13	S	Std
ICV13 - 1	ICV13	14965	3.9051	-7470A	PPB	PPB	24 Apr 2020 09:26:44	U	SMPL
ICB13 - 1	ICB13	48	-0.0014	-7470A	PPB	PPB	24 Apr 2020 09:28:49	U	SMPL
CCV89 - 1	CCV89	19430	5.0744	-7470A	PPB	PPB	24 Apr 2020 09:30:55	U	SMPL
CCB89 - 1	CCB89	58	0.0012	-7470A	PPB	PPB	24 Apr 2020 09:33:01	U	SMPL
CRA - 1	CRA	772	0.1882	-7470A	PPB	PPB	24 Apr 2020 09:35:06	U	SMPL
HighStd - 1	HighStd	37820	9.8904	-7470A	PPB	PPB	24 Apr 2020 09:37:12	U	SMPL
ChkStd - 1	ChkStd	26584	6.9479	-7470A	PPB	PPB	24 Apr 2020 09:39:18	U	SMPL
PB128445BL - 1	PBW	-70	-0.0323	-7470A	PPB	PPB	24 Apr 2020 09:41:26	U	SMPL
PB128445BS - 1	LCSW	14878	3.8823	-7470A	PPB	PPB	24 Apr 2020 09:45:54	U	SMPL
L2117-10 - 1	OK-01-042220	19	-0.0090	-7470A	PPB	PPB	24 Apr 2020 09:48:01	U	SMPL
L2117-10DUP - 1	OK-01-042220DUP	90	0.0096	-7470A	PPB	PPB	24 Apr 2020 09:50:08	U	SMPL
L2117-10MS - 1	OK-01-042220MS	14830	3.8697	-7470A	PPB	PPB	24 Apr 2020 09:54:37	U	SMPL
L2117-10MSD - 1	OK-01-042220MSD	15006	3.9158	-7470A	PPB	PPB	24 Apr 2020 09:56:45	U	SMPL
L2120-05 - 1	OR-02-042220	-12	-0.0171	-7470A	PPB	PPB	24 Apr 2020 09:58:51	U	SMPL
CCV90 - 1	CCV90	19076	4.9817	-7470A	PPB	PPB	24 Apr 2020 10:00:57	U	SMPL
CCB90 - 1	CCB90	67	0.0036	-7470A	PPB	PPB	24 Apr 2020 10:03:02	U	SMPL
L2362-01 - 1	31040	86	0.0086	-7470A	PPB	PPB	24 Apr 2020 10:05:09	U	SMPL
L2362-04 - 1	31005	109	0.0146	-7470A	PPB	PPB	24 Apr 2020 10:07:15	U	SMPL
L2364-04 - 1	30963	148	0.0248	-7470A	PPB	PPB	24 Apr 2020 10:09:20	U	SMPL
L2364-07 - 1	30964	123	0.0182	-7470A	PPB	PPB	24 Apr 2020 10:11:27	U	SMPL
L2364-10 - 1	30965	78	0.0065	-7470A	PPB	PPB	24 Apr 2020 10:13:33	U	SMPL
L2374-03 - 1	WC-1A	97	0.0114	-7470A	PPB	PPB	24 Apr 2020 10:15:40	U	SMPL
L2378-04 - 1	TP-6-S	84	0.0080	-7470A	PPB	PPB	24 Apr 2020 10:17:47	U	SMPL
L2378-08 - 1	TP-3-S	92	0.0101	-7470A	PPB	PPB	24 Apr 2020 10:19:55	U	SMPL
L2378-12 - 1	TP-2-S	98	0.0117	-7470A	PPB	PPB	24 Apr 2020 10:22:02	U	SMPL
L2383-03 - 1	ETGI-355-72-11939	134	0.0211	-7470A	PPB	PPB	24 Apr 2020 10:24:08	U	SMPL
CCV91 - 1	CCV91	19410	5.0691	-7470A	PPB	PPB	24 Apr 2020 10:26:14	U	SMPL
CCB91 - 1	CCB91	83	0.0078	-7470A	PPB	PPB	24 Apr 2020 10:28:21	U	SMPL
L2383-06 - 1	VNJ-239	127	0.0193	-7470A	PPB	PPB	24 Apr 2020 10:30:27	U	SMPL
L2397-06 - 1	TP-1-COMP	161	0.0282	-7470A	PPB	PPB	24 Apr 2020 10:32:33	U	SMPL
PB128445TB - 1	PB128424TB	108	0.0143	-7470A	PPB	PPB	24 Apr 2020 10:34:39	U	SMPL
L2117-10LX5 - 1		165	0.0292	-7470A	PPB	PPB	24 Apr 2020 10:36:46	U	SMPL
L2117-10A - 1		14968	3.9059	-7470A	PPB	PPB	24 Apr 2020 10:38:52	U	SMPL
CCV92 - 1	CCV92	19621	5.1244	-7470A	PPB	PPB	24 Apr 2020 10:43:25	U	SMPL
CCB92 - 1	CCB92	61	0.0020	-7470A	PPB	PPB	24 Apr 2020 10:45:32	U	SMPL