



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108862

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/23/2020	17:15
2	CAL2	1	Water	NA	NA	20.2	7.00	04/23/2020	17:16
3	CAL3	1	Water	NA	NA	20.2	10.04	04/23/2020	17:18
4	ICV	1	Water	NA	NA	20.2	7.01	04/23/2020	17:20
5	CCV1	1	Water	NA	NA	20.1	2.02	04/23/2020	17:22
6	L2118-06	1	Solid	20.03	20	21.5	8.45	04/23/2020	17:25
7	L2399-04	1	Solid	20.04	20	21.5	8.67	04/23/2020	17:26
8	L2399-09	1	Solid	20.05	20	21.3	8.41	04/23/2020	17:29
9	L2401-02	1	Solid	20.02	20	20.9	7.40	04/23/2020	17:32
10	L2401-02DUP	1	Solid	20.01	20	20.8	7.41	04/23/2020	17:33
11	CCV2	1	Water	NA	NA	20.1	12.02	04/23/2020	17:35

WORKLIST(Hardcopy Internal Chain)

18108862

WorkList Name : CORROSIVITY L2399

WorkList ID : 138175

Department : Wet-Chemistry

Date : 04-23-2020 14:32:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/23/2020	Solid	L2118-06	Corrosivity	Cool 4 deg C	PSEG05	J13	EO-02-042320	04/23/2020	9045D
04/23/2020	Solid	L2399-04	Corrosivity	Cool 4 deg C	PSEG01	J13	SP1-A	04/23/2020	9045D
04/23/2020	Solid	L2399-09	Corrosivity	Cool 4 deg C	PSEG01	J13	SP1-B	04/23/2020	9045D
04/23/2020	Solid	L2401-02	Corrosivity	Cool 4 deg C	PSEG01	J13	DRAINAGE-SEWALE	04/23/2020	9045D

Date/Time 04.23.2020 16:00

Received by: Jh

Relinquished by: Jh

Date/Time 04.23.2020 18:00

Received by: Jh

Relinquished by: Jh