



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108863

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP83725
Residual chlorine ICV-LCS, 0.4PPM	WP83724
Chlorine Calibration std, 0.1ppm	WP83719
Chlorine Calibration std, 0.2ppm	WP83720
Chlorine Calibration std, 0.8ppm	WP83722
Chlorine Calibration std, 0.0ppm	WP83718
Chlorine Calibration std, 1.6ppm	WP83723
Residual Chlorine LOQ 0.1ppm	WP83726
Residual Chlorine Calibration and CCV std, 0	WP83721
POTASSIUM BROMATE 0.1N 1L	W2449

Intercept: -0.0011

Slope: 1.0312

Regression: 0.999900

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/24/2020	11:04
2	CAL2	0.1	1	0.000	0.090	0.090	04/24/2020	11:04
3	CAL3	0.2	1	0.000	0.220	0.220	04/24/2020	11:05
4	CAL4	0.4	1	0.000	0.410	0.410	04/24/2020	11:06
5	CAL5	0.8	1	0.000	0.820	0.820	04/24/2020	11:07
6	CAL6	1.6	1	0.000	1.650	1.650	04/24/2020	11:08



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108863

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3800	0.3800	0.3700	04/24/2020	11:09
2	ICB				1	0.0000	0.0000	0.0000	0.0010	04/24/2020	11:10
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.3990	04/24/2020	11:11
4	CCB1				1	0.0000	0.0000	0.0000	0.0010	04/24/2020	11:12
5	LB108863BL	50	50		1	0.0000	0.0000	0.0000	0.0010	04/24/2020	11:13
6	LB108863BS	50	50	0.4	1	0.0000	0.3900	0.3900	0.3790	04/24/2020	11:14
7	L2182-07	50	50		1	0.0000	0.0600	0.0600	0.0590	04/24/2020	11:15
8	L2182-08	50	50		1	0.0000	0.1100	0.1100	0.1080	04/24/2020	11:16
9	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3890	04/24/2020	11:16
10	CCB2				1	0.0000	0.0000	0.0000	0.0010	04/24/2020	11:17