

LB 108864


INSTRUMENT ID : CV1

7471B

Linear

 μ Abs. :

54925



A= 0.0000e+000

B= 1.8215e-004

C= -1.0323e-001

Rho= 0.9997173

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	331	0.000	331				
0.2	0.200	0.147	-0.053	1376	0.0 %	1376				
2.5	2.500	2.512	0.012	14358	0.0 %	14358				
5.0	5.000	5.178	0.178	28994	0.0 %	28994				
7.5	7.500	7.504	0.004	41763	0.0 %	41763				
10.0	10.000	9.901	-0.099	54925	0.0 %	54925				

LB108864 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Apr 2020 11:37:15

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	331	-	0.00007471B	PPB	24 Apr 2020 11:57:14	S	Std
0.2 - 1	50.2	1376	-	0.20007471B	PPB	24 Apr 2020 11:59:30	S	Std
2.5 - 1	52.5	14358	-	2.50007471B	PPB	24 Apr 2020 12:01:47	S	Std
5.0 - 1	55	28994	-	5.00007471B	PPB	24 Apr 2020 12:04:04	S	Std
7.5 - 1	57.5	41763	-	7.50007471B	PPB	24 Apr 2020 12:06:21	S	Std
10.0 - 1	59.5	54925	-	10.00007471B	PPB	24 Apr 2020 12:11:11	S	Std
ICV14 - 1	ICV14	21506	3.8141	-7471B	PPB	24 Apr 2020 12:13:59	U	SMPL
ICB14 - 1	ICB14	185	-0.0695	-7471B	PPB	24 Apr 2020 12:16:13	U	SMPL
CCV93 - 1	CCV93	29069	5.1917	-7471B	PPB	24 Apr 2020 12:18:28	U	SMPL
CCB93 - 1	CCB93	181	-0.0703	-7471B	PPB	24 Apr 2020 12:20:43	U	SMPL
CRA - 1	CRA	1342	0.1412	-7471B	PPB	24 Apr 2020 12:22:59	U	SMPL
HighStd - 1	HighStd	55537	10.0129	-7471B	PPB	24 Apr 2020 12:25:14	U	SMPL
ChkStd - 1	ChkStd	38437	6.8981	-7471B	PPB	24 Apr 2020 12:27:29	U	SMPL
PB128466BL - 1	PBS	286	-0.0511	-7471B	PPB	24 Apr 2020 12:32:18	U	SMPL
PB128466BS - 1	LCSS	20083	3.5549	-7471B	PPB	24 Apr 2020 12:34:33	U	SMPL
L2182-01 - 1	LOD-MDL-SOIL-01-QT2-2020	870	0.0552	-7471B	PPB	24 Apr 2020 12:44:33	U	SMPL
L2182-02 - 1	LOQ-SOIL-02-QT2-2020	1894	0.2418	-7471B	PPB	24 Apr 2020 12:51:53	U	SMPL
L2399-01 - 1	SP1-A	2486	0.3496	-7471B	PPB	24 Apr 2020 12:54:09	U	SMPL
L2399-01DUP - 1	SP1-ADUP	2665	0.3822	-7471B	PPB	24 Apr 2020 12:56:27	U	SMPL
L2399-01MS - 1	SP1-AMS	23212	4.1249	-7471B	PPB	24 Apr 2020 12:58:42	U	SMPL
CCV94 - 1	CCV94	28593	5.1050	-7471B	PPB	24 Apr 2020 13:00:57	U	SMPL
CCB94 - 1	CCB94	159	-0.0743	-7471B	PPB	24 Apr 2020 13:03:13	U	SMPL
L2399-01MSD - 1	SP1-AMSD	23208	4.1241	-7471B	PPB	24 Apr 2020 13:05:29	U	SMPL
L2399-06 - 1	SP1-B	2093	0.2780	-7471B	PPB	24 Apr 2020 13:07:45	U	SMPL
L2401-01 - 1	DRAINAGE-SEWALE	2024	0.2654	-7471B	PPB	24 Apr 2020 13:10:01	U	SMPL
L2401-05 - 1	CONCRETE-DEMO-1	606	0.0072	-7471B	PPB	24 Apr 2020 13:12:16	U	SMPL
L2401-06 - 1	CONCRETE-DEMO-2	811	0.0445	-7471B	PPB	24 Apr 2020 13:14:32	U	SMPL
L2399-01LX5 - 1		867	0.0547	-7471B	PPB	24 Apr 2020 13:16:49	U	SMPL
L2399-01A - 1		24425	4.3458	-7471B	PPB	24 Apr 2020 13:19:05	U	SMPL
CCV95 - 1	CCV95	28650	5.1154	-7471B	PPB	24 Apr 2020 13:21:22	U	SMPL
CCB95 - 1	CCB95	179	-0.0706	-7471B	PPB	24 Apr 2020 13:23:40	U	SMPL