

LB108869


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

39088



A= 0.0000e+000

B= 2.5881e-004

C= -3.6767e-002

Rho= 0.9996626

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	166	0.000	166				
0.2	0.200	0.210	0.010	953	0.0 %	953				
2.5	2.500	2.468	-0.032	9676	0.0 %	9676				
5.0	5.000	5.119	0.119	19922	0.0 %	19922				
7.5	7.500	7.317	-0.183	28415	0.0 %	28415				
10.0	10.000	10.080	0.080	39088	0.0 %	39088				

LB108869 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Apr 2020 14:16:20

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	166	-	0.0000	7470A	PPB	24 Apr 2020 14:24:34	S	Std
0.2 - 1	50.2	953	-	0.2000	7470A	PPB	24 Apr 2020 14:26:41	S	Std
2.5 - 1	52.5	9676	-	2.5000	7470A	PPB	24 Apr 2020 14:28:49	S	Std
5.0 - 1	55	19922	-	5.0000	7470A	PPB	24 Apr 2020 14:30:56	S	Std
7.5 - 1	57.5	28415	-	7.5000	7470A	PPB	24 Apr 2020 14:33:03	S	Std
10.0 - 1	510	39088	-	10.0000	7470A	PPB	24 Apr 2020 14:35:09	S	Std
ICV15 - 1	ICV15	14886	3.8159	-	7470A	PPB	24 Apr 2020 14:37:44	U	SMPL
ICB15 - 1	ICB15	46	-0.0249	-	7470A	PPB	24 Apr 2020 14:39:49	U	SMPL
CCV96 - 1	CCV96	20193	5.1894	-	7470A	PPB	24 Apr 2020 14:41:55	U	SMPL
CCB96 - 1	CCB96	37	-0.0272	-	7470A	PPB	24 Apr 2020 14:44:00	U	SMPL
CRA - 1	CRA	887	0.1928	-	7470A	PPB	24 Apr 2020 14:46:06	U	SMPL
HighStd - 1	HighStd	39740	10.2485	-	7470A	PPB	24 Apr 2020 14:48:11	U	SMPL
ChkStd - 1	ChkStd	25038	6.4434	-	7470A	PPB	24 Apr 2020 14:50:18	U	SMPL
PB128474BL - 1	PBW	103	-0.0101	-	7470A	PPB	24 Apr 2020 14:52:26	U	SMPL
PB128474BS - 1	LCSW	15651	4.0139	-	7470A	PPB	24 Apr 2020 14:54:33	U	SMPL
L2118-06 - 1	EO-02-042320	45	-0.0251	-	7470A	PPB	24 Apr 2020 14:56:40	U	SMPL
L2118-06DUP - 1	EO-02-042320DUP	132	-0.0026	-	7470A	PPB	24 Apr 2020 14:58:47	U	SMPL
L2118-06MS - 1	EO-02-042320MS	13835	3.5439	-	7470A	PPB	24 Apr 2020 15:00:55	U	SMPL
L2118-06MSD - 1	EO-02-042320MSD	14066	3.6037	-	7470A	PPB	24 Apr 2020 15:03:02	U	SMPL
L2314-19 - 1	WB-2-1	156	0.0036	-	7470A	PPB	24 Apr 2020 15:05:08	U	SMPL
CCV97 - 1	CCV97	20287	5.2138	-	7470A	PPB	24 Apr 2020 15:07:14	U	SMPL
CCB97 - 1	CCB97	47	-0.0246	-	7470A	PPB	24 Apr 2020 15:09:20	U	SMPL
L2314-20 - 1	WB-2-2	195	0.0137	-	7470A	PPB	24 Apr 2020 15:11:25	U	SMPL
L2314-21 - 1	WB-3-1	108	-0.0088	-	7470A	PPB	24 Apr 2020 15:13:31	U	SMPL
L2314-22 - 1	WB-3-2	132	-0.0026	-	7470A	PPB	24 Apr 2020 15:15:37	U	SMPL
L2314-23 - 1	WB-4-1	81	-0.0158	-	7470A	PPB	24 Apr 2020 15:17:44	U	SMPL
L2314-26 - 1	WB-4-2	267	0.0323	-	7470A	PPB	24 Apr 2020 15:19:50	U	SMPL
L2314-27 - 1	WB-5-1	210	0.0176	-	7470A	PPB	24 Apr 2020 15:21:57	U	SMPL
L2314-28 - 1	WB-5-2	231	0.0230	-	7470A	PPB	24 Apr 2020 15:24:04	U	SMPL
L2314-29 - 1	DUP-2020-04-15	250	0.0279	-	7470A	PPB	24 Apr 2020 15:26:12	U	SMPL
L2314-30 - 1	WB-6-1	245	0.0266	-	7470A	PPB	24 Apr 2020 15:28:20	U	SMPL
L2314-31 - 1	WB-6-2	254	0.0290	-	7470A	PPB	24 Apr 2020 15:30:26	U	SMPL
CCV98 - 1	CCV98	20442	5.2539	-	7470A	PPB	24 Apr 2020 15:32:32	U	SMPL
CCB98 - 1	CCB98	73	-0.0179	-	7470A	PPB	24 Apr 2020 15:34:38	U	SMPL
L2314-32 - 1	WB-7-1	127	-0.0039	-	7470A	PPB	24 Apr 2020 15:36:44	U	SMPL
L2314-33 - 1	WB-7-2	154	0.0031	-	7470A	PPB	24 Apr 2020 15:38:51	U	SMPL
L2314-34 - 1	WB-8-1	201	0.0153	-	7470A	PPB	24 Apr 2020 15:40:57	U	SMPL
L2314-35 - 1	WB-8-2	150	0.0021	-	7470A	PPB	24 Apr 2020 15:43:03	U	SMPL
L2399-04 - 1	SP1-A	145	0.0008	-	7470A	PPB	24 Apr 2020 15:45:10	U	SMPL
L2399-09 - 1	SP1-B	169	0.0070	-	7470A	PPB	24 Apr 2020 15:47:17	U	SMPL
L2401-02 - 1	DRAINAGE-SEWALE	153	0.0028	-	7470A	PPB	24 Apr 2020 15:49:24	U	SMPL
PB128474TB - 1	PB128408TB	178	0.0093	-	7470A	PPB	24 Apr 2020 15:51:32	U	SMPL
L2118-06LX5 - 1		148	0.0015	-	7470A	PPB	24 Apr 2020 15:53:39	U	SMPL
L2118-06A - 1		14969	3.8374	-	7470A	PPB	24 Apr 2020 15:55:45	U	SMPL
CCV99 - 1	CCV99	20467	5.2604	-	7470A	PPB	24 Apr 2020 15:57:51	U	SMPL
CCB99 - 1	CCB99	58	-0.0218	-	7470A	PPB	24 Apr 2020 15:59:58	U	SMPL