



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108876

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/24/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/24/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/24/2020	17:19
4	ICV	1	Water	NA	NA	20.2	6.99	04/24/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	04/24/2020	17:22
6	L2420-01	1	Solid	20.02	20	21.6	5.62	04/24/2020	17:25
7	L2420-03	1	Solid	20.04	20	21.6	9.48	04/24/2020	17:29
8	L2420-12	1	Solid	20.01	20	21.4	5.56	04/24/2020	17:32
9	L2422-02	1	Solid	20.03	20	21.5	7.10	04/24/2020	17:35
10	L2422-02DUP	1	Solid	20.04	20	21.6	7.11	04/24/2020	17:36
11	CCV2	1	Water	NA	NA	20.1	12.02	04/24/2020	17:38

108876

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2382 WorkList ID : 138232 Department : Wet-Chemistry Date : 04-24-2020 16:11:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2020	Solid	L2420-01	Corrosivity	Cool 4 deg C	PSEG01	J21	DRUMS-C-1-4	04/24/2020	9045D
04/24/2020	Solid	L2420-03	Corrosivity	Cool 4 deg C	PSEG01	J21	DRUMS-D-1-2	04/24/2020	9045D
04/24/2020	Solid	L2420-12	Corrosivity	Cool 4 deg C	PSEG01	J21	DRUMS-B-1-3	04/24/2020	9045D
04/24/2020	Solid	L2422-02	Corrosivity	Cool 4 deg C	PSEG01	J21	DUCK-ISLAND-ROLLOFF	04/24/2020	9045D

Date/Time 04-24-2020 16:12
 Received by: JL
 Relinquished by: JL

Date/Time 04-24-2020 18:20
 Received by: JL
 Relinquished by: JL