



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108877

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	04/24/2020	17:15
2	CAL2	1	Water	NA	NA	20.1	7.01	04/24/2020	17:16
3	CAL3	1	Water	NA	NA	20.1	10.03	04/24/2020	17:19
4	ICV	1	Water	NA	NA	20.2	6.99	04/24/2020	17:20
5	CCV1	1	Water	NA	NA	20.2	2.01	04/24/2020	17:22
6	L2420-05	1	Water	NA	NA	21.3	5.54	04/24/2020	17:40
7	L2420-05DUP	1	Water	NA	NA	21.4	5.55	04/24/2020	17:41
8	CCV2	1	Water	NA	NA	20.2	12.02	04/24/2020	17:43

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2420 W WorkList ID : 138233 Department : Wet-Chemistry Date : 04-24-2020 16:22:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/24/2020	Water	L2420-05	Corrosivity	Cool 4 deg C	PSEG01	J21	DRUMS-A-1-2	04/24/2020	9040C

Date/Time 04.24.2020 16:12
 Received by: JO
 Relinquished by: JO

Date/Time 04.24.2020 18:20
 Received by: JO
 Relinquished by: JO