

LB108888


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38017



A= 0.0000e+000

B= 2.6465e-004

C= -3.6372e-002

Rho= 0.9999031

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.021	-0.021	57	0.000	57				
0.2	0.200	0.181	-0.019	823	0.0 %	823				
2.5	2.500	2.542	0.042	9743	0.0 %	9743				
5.0	5.000	5.065	0.065	19276	0.0 %	19276				
7.5	7.500	7.408	-0.092	28130	0.0 %	28130				
10.0	10.000	10.025	0.025	38017	0.0 %	38017				

LB108888 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 27 Apr 2020 15:14:25					
Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7470A	PPB	27 Apr 2020 15:15:38	S	Std
0.2 - 1	50-2	823	-	0.2000	7470A	PPB	27 Apr 2020 15:17:46	S	Std
2.5 - 1	52-5	9743	-	2.5000	7470A	PPB	27 Apr 2020 15:19:54	S	Std
5.0 - 1	55	19276	-	5.0000	7470A	PPB	27 Apr 2020 15:22:02	S	Std
7.5 - 1	57-5	28130	-	7.5000	7470A	PPB	27 Apr 2020 15:24:09	S	Std
10.0 - 1	510	38017	-	10.0000	7470A	PPB	27 Apr 2020 15:26:15	S	Std
ICV18 - 1	ICV18	14818	3.8852	-	7470A	PPB	27 Apr 2020 15:28:45	U	SMPL
ICB18 - 1	ICB18	-100	-0.0628	-	7470A	PPB	27 Apr 2020 15:30:51	U	SMPL
CCV07 - 1	CCV07	19221	5.0504	-	7470A	PPB	27 Apr 2020 15:32:56	U	SMPL
CCB07 - 1	CCB07	-107	-0.0647	-	7470A	PPB	27 Apr 2020 15:35:02	U	SMPL
CRA - 1	CRA	778	0.1695	-	7470A	PPB	27 Apr 2020 15:37:08	U	SMPL
HighStd - 1	HighStd	37992	10.0181	-	7470A	PPB	27 Apr 2020 15:39:14	U	SMPL
ChkStd - 1	ChkStd	26237	6.9072	-	7470A	PPB	27 Apr 2020 15:41:20	U	SMPL
PB128507BL - 1	PBW	-137	-0.0726	-	7470A	PPB	27 Apr 2020 15:43:28	U	SMPL
PB128507BS - 1	LCSW	13794	3.6142	-	7470A	PPB	27 Apr 2020 15:45:36	U	SMPL
L2382-12 - 1	19972	580	0.1171	-	7470A	PPB	27 Apr 2020 15:47:43	U	SMPL
L2422-02 - 1	DUCK-ISLAND-ROLLOFF	-15	-0.0403	-	7470A	PPB	27 Apr 2020 15:49:50	U	SMPL
L2422-02DUP - 1	DUCK-ISLAND-ROLLOFFDUP	27	-0.0292	-	7470A	PPB	27 Apr 2020 15:51:58	U	SMPL
L2422-02MS - 1	DUCK-ISLAND-ROLLOFFMS	13300	3.4834	-	7470A	PPB	27 Apr 2020 15:54:06	U	SMPL
L2422-02MSD - 1	DUCK-ISLAND-ROLLOFFMSD	12995	3.4027	-	7470A	PPB	27 Apr 2020 15:56:11	U	SMPL
CCV08 - 1	CCV08	18699	4.9123	-	7470A	PPB	27 Apr 2020 15:58:17	U	SMPL
CCB08 - 1	CCB08	-102	-0.0634	-	7470A	PPB	27 Apr 2020 16:00:22	U	SMPL
L2422-02LX5 - 1		27	-0.0292	-	7470A	PPB	27 Apr 2020 16:07:13	U	SMPL
L2422-02A - 1		14820	3.8857	-	7470A	PPB	27 Apr 2020 16:09:19	U	SMPL
PB128507TB - 1	PB128492TB	-47	-0.0488	-	7470A	PPB	27 Apr 2020 16:11:25	U	SMPL
CCV09 - 1	CCV09	19015	4.9959	-	7470A	PPB	27 Apr 2020 16:13:31	U	SMPL
CCB09 - 1	CCB09	-97	-0.0620	-	7470A	PPB	27 Apr 2020 16:15:37	U	SMPL