



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108899

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP83825
Residual chlorine ICV-LCS, 0.4PPM	WP83821
Chlorine Calibration std, 0.1ppm	WP83819
Chlorine Calibration std, 0.2ppm	WP83820
Chlorine Calibration std, 0.8ppm	WP83822
Chlorine Calibration std, 0.0ppm	WP83818
Chlorine Calibration std, 1.6ppm	WP83823
Residual Chlorine Calibration and CCV std, 0	WP83824
POTASSIUM BROMATE 0.1N 1L	W2449

Intercept: 0.0020

Slope: 1.0316

Regression: 0.999980

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	04/28/2020	11:10
2	CAL2	0.1	1	0.000	0.100	0.100	04/28/2020	11:11
3	CAL3	0.2	1	0.000	0.210	0.210	04/28/2020	11:12
4	CAL4	0.4	1	0.000	0.420	0.420	04/28/2020	11:13
5	CAL5	0.8	1	0.000	0.830	0.830	04/28/2020	11:14
6	CAL6	1.6	1	0.000	1.650	1.650	04/28/2020	11:15



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB108899

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.4260	04/28/2020	11:18
2	ICB				1	0.0000	0.0000	0.0000	-0.0010	04/28/2020	11:19
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.3970	04/28/2020	11:20
4	CCB1				1	0.0000	0.0000	0.0000	-0.0010	04/28/2020	11:22
5	LB108899BL	50	50		1	0.0000	0.0000	0.0000	-0.0010	04/28/2020	11:24
6	LB108899BS	50	50	0.4	1	0.0000	0.4400	0.4400	0.4260	04/28/2020	11:25
7	L2182-09	50	50		1	0.0000	0.0500	0.0500	0.0470	04/28/2020	11:27
8	CCV2			0.4	1	0.0000	0.4300	0.4300	0.4160	04/28/2020	11:28
9	CCB2				1	0.0000	0.0000	0.0000	-0.0010	04/28/2020	11:30