

LB108912


7471B

INSTRUMENT ID: CY1

Linear

 μ Abs.:

55883



A= 0.0000e+000

B= 1.7702e-004

C= -3.8974e-002

Rho= 0.9996995

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	26	0.000	26				
0.2	0.200	0.165	-0.035	1152	0.0 %	1152				
2.5	2.500	2.506	0.006	14377	0.0 %	14377				
5.0	5.000	5.066	0.066	28841	0.0 %	28841				
7.5	7.500	7.644	0.144	43402	0.0 %	43402				
10.0	10.000	9.853	-0.147	55883	0.0 %	55883				

LB108912 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Apr 2020 10:54:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7471B	PPB	29 Apr 2020 11:02:19	S	Std
0.2 - 1	50.2	1152	-	0.2000	7471B	PPB	29 Apr 2020 11:04:36	S	Std
2.5 - 1	52.5	14377	-	2.5000	7471B	PPB	29 Apr 2020 11:06:53	S	Std
5.0 - 1	55	28841	-	5.0000	7471B	PPB	29 Apr 2020 11:09:10	S	Std
7.5 - 1	57.5	43402	-	7.5000	7471B	PPB	29 Apr 2020 11:11:26	S	Std
10.0 - 1	51.0	55883	-	10.0000	7471B	PPB	29 Apr 2020 11:13:43	S	Std
ICV20 - 1	ICV20	21150	3.7049	-	7471B	PPB	29 Apr 2020 11:16:39	U	SMPL
ICB20 - 1	ICB20	-73	-0.0519	-	7471B	PPB	29 Apr 2020 11:18:54	U	SMPL
CCV13 - 1	CCV13	28383	4.9853	-	7471B	PPB	29 Apr 2020 11:21:09	U	SMPL
CCB13 - 1	CCB13	-92	-0.0553	-	7471B	PPB	29 Apr 2020 11:23:23	U	SMPL
CRA - 1	CRA	1117	0.1588	-	7471B	PPB	29 Apr 2020 11:25:39	U	SMPL
HighStd - 1	HighStd	55731	9.8263	-	7471B	PPB	29 Apr 2020 11:27:54	U	SMPL
ChkStd - 1	ChkStd	40150	7.0682	-	7471B	PPB	29 Apr 2020 11:30:09	U	SMPL
PB128533BL - 1	PBS	-122	-0.0606	-	7471B	PPB	29 Apr 2020 11:32:26	U	SMPL
PB128533BS - 1	LCSS	20338	3.5612	-	7471B	PPB	29 Apr 2020 11:38:03	U	SMPL
L2436-01 - 1	HEALTH	2102	0.3331	-	7471B	PPB	29 Apr 2020 11:45:17	U	SMPL
L2436-01DUP - 1	HEALTHDUP	2037	0.3216	-	7471B	PPB	29 Apr 2020 11:47:33	U	SMPL
L2436-01MS - 1	HEALTHMS	21703	3.8028	-	7471B	PPB	29 Apr 2020 11:49:49	U	SMPL
L2436-01MSD - 1	HEALTHMSD	21794	3.8189	-	7471B	PPB	29 Apr 2020 11:52:06	U	SMPL
CCV14 - 1	CCV14	28224	4.9571	-	7471B	PPB	29 Apr 2020 11:54:21	U	SMPL
CCB14 - 1	CCB14	-62	-0.0499	-	7471B	PPB	29 Apr 2020 11:56:37	U	SMPL
L2436-01LX5 - 1		437	0.0384	-	7471B	PPB	29 Apr 2020 11:58:52	U	SMPL
L2436-01A - 1		23588	4.1365	-	7471B	PPB	29 Apr 2020 12:01:07	U	SMPL
CCV15 - 1	CCV15	28722	5.0453	-	7471B	PPB	29 Apr 2020 12:03:23	U	SMPL
CCB15 - 1	CCB15	-74	-0.0521	-	7471B	PPB	29 Apr 2020 12:05:39	U	SMPL