


LB108913

245.1

INSTRUMENT ID: CV4

 μ Abs.:

38085



Linear

A= 0.0000e+000

B= 2.6344e-004

C= -6.5680e-004

Rho= 0.9999797

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	71	0.000	71				
0.05	0.050	0.056	0.006	215	0.0 %	215				
0.2	0.200	0.207	0.007	789	0.0 %	789				
2.5	2.500	2.467	-0.033	9369	0.0 %	9369				
5.0	5.000	5.006	0.006	19005	0.0 %	19005				
7.5	7.500	7.463	-0.037	28332	0.0 %	28332				
10.0	10.000	10.032	0.032	38085	0.0 %	38085				

LB108913 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 29 Apr 2020 12:19:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	71	-	0.0000	245.1	PPB	29 Apr 2020 12:20:41	S	Std
0.05 - 1	50.05	215	-	0.0500	245.1	PPB	29 Apr 2020 12:22:57	S	Std
0.2 - 1	50.2	789	-	0.2000	245.1	PPB	29 Apr 2020 12:25:14	S	Std
2.5 - 1	52.5	9369	-	2.5000	245.1	PPB	29 Apr 2020 12:27:31	S	Std
5.0 - 1	55	19005	-	5.0000	245.1	PPB	29 Apr 2020 12:29:47	S	Std
7.5 - 1	57.5	28332	-	7.5000	245.1	PPB	29 Apr 2020 12:32:03	S	Std
10.0 - 1	510	38085	-	10.0000	245.1	PPB	29 Apr 2020 12:45:25	S	Std
ICV21 - 1	ICV21	14748	3.8845	-245.1		PPB	29 Apr 2020 12:53:12	U	SMPL
ICB21 - 1	ICB21	-83	-0.0225	-245.1		PPB	29 Apr 2020 12:55:26	U	SMPL
CCV16 - 1	CCV16	19486	5.1327	-245.1		PPB	29 Apr 2020 12:57:41	U	SMPL
CCB16 - 1	CCB16	-100	-0.0270	-245.1		PPB	29 Apr 2020 12:59:56	U	SMPL
CRA - 1	CRA	197	0.0512	-245.1		PPB	29 Apr 2020 13:02:11	U	SMPL
HighStd - 1	HighStd	37764	9.9478	-245.1		PPB	29 Apr 2020 13:04:26	U	SMPL
ChkStd - 1	ChkStd	26382	6.9493	-245.1		PPB	29 Apr 2020 13:06:42	U	SMPL
PB128540BL - 1	PBW	-143	-0.0383	-245.1		PPB	29 Apr 2020 13:08:59	U	SMPL
PB128540BS - 1	LCSW	14119	3.7188	-245.1		PPB	29 Apr 2020 13:11:16	U	SMPL
L2430-01 - 1	37-ST-2020	-11	-0.0036	-245.1		PPB	29 Apr 2020 13:13:32	U	SMPL
L2430-01DUP - 1	37-ST-2020DUP	60	0.0151	-245.1		PPB	29 Apr 2020 13:15:49	U	SMPL
L2430-01MS - 1	37-ST-2020MS	15848	4.1743	-245.1		PPB	29 Apr 2020 13:18:05	U	SMPL
L2430-01MSD - 1	37-ST-2020MSD	15842	4.1727	-245.1		PPB	29 Apr 2020 13:20:23	U	SMPL
CCV17 - 1	CCV17	19281	5.0787	-245.1		PPB	29 Apr 2020 13:22:38	U	SMPL
CCB17 - 1	CCB17	-103	-0.0278	-245.1		PPB	29 Apr 2020 13:24:54	U	SMPL
L2430-01LX5 - 1		22	0.0051	-245.1		PPB	29 Apr 2020 13:27:09	U	SMPL
L2430-01A - 1		14540	3.8297	-245.1		PPB	29 Apr 2020 13:29:24	U	SMPL
CCV18 - 1	CCV18	18900	4.9783	-245.1		PPB	29 Apr 2020 13:31:39	U	SMPL
CCB18 - 1	CCB18	-82	-0.0223	-245.1		PPB	29 Apr 2020 13:33:55	U	SMPL