


LB108918

7470A

INSTRUMENT ID: CYJ

 μ Abs.:

38495



A= 0.0000e+000

B= 2.6190e-004

C= -4.9058e-002

Rho= 0.9997693

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	56	0.000	56				
0.2	0.200	0.151	-0.049	763	0.0 %	763				
2.5	2.500	2.619	0.119	10187	0.0 %	10187				
5.0	5.000	5.057	0.057	19496	0.0 %	19496				
7.5	7.500	7.375	-0.125	28347	0.0 %	28347				
10.0	10.000	10.033	0.033	38495	0.0 %	38495				

LB108918 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Apr 2020 14:17:12

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	56	-	0.00007470A	PPB	29 Apr 2020 14:18:30	S	Std
0.2 - 1	50.2	763	-	0.20007470A	PPB	29 Apr 2020 14:20:37	S	Std
2.5 - 1	52.5	10187	-	2.50007470A	PPB	29 Apr 2020 14:22:46	S	Std
5.0 - 1	55	19496	-	5.00007470A	PPB	29 Apr 2020 14:24:53	S	Std
7.5 - 1	57.5	28347	-	7.50007470A	PPB	29 Apr 2020 14:27:00	S	Std
10.0 - 1	510	38495	-	10.00007470A	PPB	29 Apr 2020 14:29:06	S	Std
ICV22 - 1	ICV22	15628	4.0439	-7470A	PPB	29 Apr 2020 14:34:20	U	SMPL
ICB22 - 1	ICB22	-78	-0.0695	-7470A	PPB	29 Apr 2020 14:36:26	U	SMPL
CCV19 - 1	CCV19	19245	4.9912	-7470A	PPB	29 Apr 2020 14:38:31	U	SMPL
CCB19 - 1	CCB19	-88	-0.0721	-7470A	PPB	29 Apr 2020 14:40:37	U	SMPL
CRA - 1	CRA	726	0.1411	-7470A	PPB	29 Apr 2020 14:42:43	U	SMPL
HighStd - 1	HighStd	37892	9.8748	-7470A	PPB	29 Apr 2020 14:44:49	U	SMPL
ChkStd - 1	ChkStd	26626	6.9243	-7470A	PPB	29 Apr 2020 14:46:55	U	SMPL
PB128557BL - 1	PBW	-137	-0.0849	-7470A	PPB	29 Apr 2020 14:49:03	U	SMPL
PB128557BS - 1	LCSW	14073	3.6367	-7470A	PPB	29 Apr 2020 14:51:11	U	SMPL
L2436-04 - 1	HEALTH	-42	-0.0601	-7470A	PPB	29 Apr 2020 14:53:18	U	SMPL
L2436-04DUP - 1	HEALTHDUP	5	-0.0477	-7470A	PPB	29 Apr 2020 14:55:25	U	SMPL
L2436-04MS - 1	HEALTHMS	15514	4.0141	-7470A	PPB	29 Apr 2020 14:57:33	U	SMPL
L2436-04MSD - 1	HEALTHMSD	15413	3.9876	-7470A	PPB	29 Apr 2020 14:59:41	U	SMPL
L2439-01 - 1	S-1	-77	-0.0692	-7470A	PPB	29 Apr 2020 15:01:46	U	SMPL
CCV20 - 1	CCV20	19290	5.0030	-7470A	PPB	29 Apr 2020 15:03:52	U	SMPL
CCB20 - 1	CCB20	-101	-0.0755	-7470A	PPB	29 Apr 2020 15:05:58	U	SMPL
L2436-04LX5 - 1		17	-0.0446	-7470A	PPB	29 Apr 2020 15:13:34	U	SMPL
L2436-04A - 1		15321	3.9635	-7470A	PPB	29 Apr 2020 15:15:40	U	SMPL
PB128557TB - 1	PB128534TB	-64	-0.0658	-7470A	PPB	29 Apr 2020 15:17:47	U	SMPL
CCV21 - 1	CCV21	18801	4.8749	-7470A	PPB	29 Apr 2020 15:19:53	U	SMPL
CCB21 - 1	CCB21	-90	-0.0726	-7470A	PPB	29 Apr 2020 15:22:00	U	SMPL