



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108924

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	04/30/2020	10:05
2	L2430-01		1	9	100.00	>212.0	>212.0	04/30/2020	10:35
3	L2430-01DUP		1	9	100.00	>212.0	>212.0	04/30/2020	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 108924

WorkList Name : fp42920-1 WorkList ID : 138314 Department : Wet-Chemistry Date : 04-29-2020 14:47:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/11/2020	Water	L2430-01	H	Flash Point	Cool 4 deg C	DIST01	37-ST-2020	04/28/2020	1010A

Date/Time 04/30/20 09:30
 Received by: KS
 Relinquished by: JS

Date/Time 04/30/20 11:15
 Received by: JP
 Relinquished by: KS