



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/30/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:15
In Date: 04/29/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 04/30/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108927

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2439-01	S-1	1	1.14	9.55	7.24	72.5	
L2441-01	RECOVERED-CRSH-AGGREGAT E-A	2	1.14	9.75	9.22	93.8	
L2441-02	RECOVERED-CRSH-AGGREGAT E-B	3	1.13	9.52	8.98	93.6	
L2441-03	RECOVERED-CRSH-AGGREGAT E-C	4	1.14	9.59	9.00	93.0	
L2443-01	BP-VPB180-GW-658-660	5	1.17	9.97	2.51	15.2	sludge sample
L2443-02	BP-VPB180-GW-678-680	6	1.15	9.95	2.46	14.9	sludge sample
L2443-03	BP-VPB180-GW-708-710	7	1.14	9.92	2.14	11.4	sludge sample
L2443-04	BP-VPB180-GW-718-720	8	1.12	9.97	2.21	12.3	sludge sample
L2445-01	EOD-PROJECT-EAST	9	1.15	9.94	9.26	92.3	
L2445-02	EOD-PROJECT-EAST	10	1.11	9.88	8.85	88.3	
L2445-03	EOD-PROJECT-WEST	11	1.11	9.65	8.66	88.4	
L2445-04	EOD-PROJECT-WEST	12	1.11	9.78	8.9	89.9	
L2447-01	PT-AN-SOIL	13	1.00	2.00	2.00	100.0	
L2447-02	PT-CORR-SOIL	14	1.00	2.00	2.00	100.0	
L2447-03	PT-CORR-SOIL	15	1.00	2.00	2.00	100.0	
L2447-04	PT-CN-SOIL	16	1.00	2.00	2.00	100.0	
L2447-05	PT-CN-SOIL	17	1.00	2.00	2.00	100.0	
L2447-06	PT-FP-SOIL	18	1.00	2.00	2.00	100.0	
L2447-07	PT-CR6-SOIL	19	1.00	2.00	2.00	100.0	
L2447-08	PT-NUT-SOIL	20	1.00	2.00	2.00	100.0	
L2447-09	PT-NUT-SOIL	21	1.00	2.00	2.00	100.0	
L2447-10	PT-OCR-SOIL	22	1.00	2.00	2.00	100.0	
L2447-11	PT-MET-SOIL	23	1.00	2.00	2.00	100.0	
L2447-12	PT-BNA-SOIL	24	1.00	2.00	2.00	100.0	
L2447-13	PT-PAH-SOIL	25	1.00	2.00	2.00	100.0	
L2447-14	PT-DIES-SOIL	26	1.00	2.00	2.00	100.0	
L2447-15	PT-GAS-SOIL	27	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 4/30/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:15
In Date: 04/29/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:36
Out Date: 04/30/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB108927

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L2447-16	PT-NJEPH-SOIL	28	1.00	2.00	2.00	100.0	
L2447-17	PT-HERB-SOIL	29	1.00	2.00	2.00	100.0	
L2447-18	PT-PCB-SOIL	30	1.00	2.00	2.00	100.0	
L2447-19	PT-PEST-SOIL	31	1.00	2.00	2.00	100.0	
L2447-20	PT-CHLR-SOIL	32	1.00	2.00	2.00	100.0	
L2447-21	PT-TXP-SOIL	33	1.00	2.00	2.00	100.0	
L2447-22	PT-VOA-SOIL	34	1.00	2.00	2.00	100.0	
L2447-23	PT-SOL-SOIL	35	1.18	9.4	8.15	84.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$