



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108930

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/29/2020	14:25
2	CAL2	1	Water	NA	NA	20.2	7.00	04/29/2020	14:26
3	CAL3	1	Water	NA	NA	20.2	10.04	04/29/2020	14:27
4	ICV	1	Water	NA	NA	20.2	7.01	04/29/2020	14:30
5	CCV1	1	Water	NA	NA	20.1	2.01	04/29/2020	14:32
6	L2436-04	1	Solid	20.03	20	21.4	6.32	04/29/2020	14:35
7	L2436-04DUP	1	Solid	20.04	20	21.5	6.33	04/29/2020	14:36
8	L2447-02	1	Solid	20.02	20	22.9	9.13	04/29/2020	14:40
9	CCV2	1	Water	NA	NA	20.1	12.02	04/29/2020	14:42

18108930

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2436 WorkList ID : 138317 Department : Wet-Chemistry Date : 04-29-2020 14:15:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/28/2020	Solid	L2436-04	Corrosivity	Cool 4 deg C	PSEG01	K11	HEALTH	04/28/2020	9045D

Date/Time 04.29.2020 14:20
 Received by: JP
 Relinquished by: JP

Date/Time 04.29.2020 15:35
 Received by: JP
 Relinquished by: JP