


LB108943

7470A

INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	63	0.000	63				
0.2	0.200	0.170	-0.030	832	0.0 %	832				
2.5	2.500	2.504	0.004	9896	0.0 %	9896				
5.0	5.000	5.126	0.126	20079	0.0 %	20079				
7.5	7.500	7.473	-0.027	29193	0.0 %	29193				
10.0	10.000	9.955	-0.045	38834	0.0 %	38834				

LB108943 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Apr 2020 12:35:42

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	63	-	0.0000	7470A	PPB	30 Apr 2020 13:18:18	S	Std
0.2 - 1	50.2	832	-	0.2000	7470A	PPB	30 Apr 2020 13:20:26	S	Std
2.5 - 1	52.5	9896	-	2.5000	7470A	PPB	30 Apr 2020 13:22:33	S	Std
5.0 - 1	55	20079	-	5.0000	7470A	PPB	30 Apr 2020 13:24:41	S	Std
7.5 - 1	57.5	29193	-	7.5000	7470A	PPB	30 Apr 2020 13:26:48	S	Std
10.0 - 1	510	38834	-	10.0000	7470A	PPB	30 Apr 2020 13:28:54	S	Std
ICV24 - 1	ICV24	15387	3.9178	-	7470A	PPB	30 Apr 2020 13:31:33	U	SMPL
ICB24 - 1	ICB24	-72	-0.0626	-	7470A	PPB	30 Apr 2020 13:33:39	U	SMPL
CCV25 - 1	CCV25	19836	5.0634	-	7470A	PPB	30 Apr 2020 13:35:44	U	SMPL
CCB25 - 1	CCB25	-88	-0.0667	-	7470A	PPB	30 Apr 2020 13:37:50	U	SMPL
CRA - 1	CRA	778	0.1562	-	7470A	PPB	30 Apr 2020 13:39:55	U	SMPL
HighStd - 1	HighStd	38819	9.9512	-	7470A	PPB	30 Apr 2020 13:42:01	U	SMPL
ChkStd - 1	ChkStd	26592	6.8030	-	7470A	PPB	30 Apr 2020 13:44:07	U	SMPL
PB128583BL - 1	PBW	-75	-0.0634	-	7470A	PPB	30 Apr 2020 13:46:15	U	SMPL
PB128583BS - 1	LCSW	15089	3.8411	-	7470A	PPB	30 Apr 2020 13:48:22	U	SMPL
L2182-07 - 1	LOD-MDL-WATER-01-QT2-2020	330	0.0409	-	7470A	PPB	30 Apr 2020 13:57:41	U	SMPL
L2182-08 - 1	LOQ-WATER-02-QT2-2020	825	0.1683	-	7470A	PPB	30 Apr 2020 13:59:48	U	SMPL
L2445-05 - 1	B-1511-T	110	-0.0158	-	7470A	PPB	30 Apr 2020 14:01:56	U	SMPL
L2445-05DUP - 1	B-1511-TDUP	129	-0.0109	-	7470A	PPB	30 Apr 2020 14:04:03	U	SMPL
L2445-05MS - 1	B-1511-TMS	13802	3.5097	-	7470A	PPB	30 Apr 2020 14:06:09	U	SMPL
CCV26 - 1	CCV26	20021	5.1110	-	7470A	PPB	30 Apr 2020 14:08:15	U	SMPL
CCB26 - 1	CCB26	-83	-0.0654	-	7470A	PPB	30 Apr 2020 14:10:21	U	SMPL
L2445-05MSD - 1	B-1511-TMSD	13069	3.3210	-	7470A	PPB	30 Apr 2020 14:12:28	U	SMPL
L2445-06 - 1	B-1511-B	46	-0.0322	-	7470A	PPB	30 Apr 2020 14:14:33	U	SMPL
L2445-07 - 1	FIELD-BLANK-4-29-2020	43	-0.0330	-	7470A	PPB	30 Apr 2020 14:16:38	U	SMPL
L2445-05LX5 - 1		52	-0.0307	-	7470A	PPB	30 Apr 2020 14:18:44	U	SMPL
L2445-05A - 1		15200	3.8697	-	7470A	PPB	30 Apr 2020 14:20:51	U	SMPL
CCV27 - 1	CCV27	19778	5.0485	-	7470A	PPB	30 Apr 2020 14:22:58	U	SMPL
CCB27 - 1	CCB27	-85	-0.0660	-	7470A	PPB	30 Apr 2020 14:25:05	U	SMPL