

Analysis Method: SM4500 S E

ANALYST: ketankumar

Parameter: Sulfide

SUPERVISOR REVIEW BY: apatel

Run Number: LB108951

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP83473
IODINE SOLUTION .025N 1L	W2669
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	T1 (mL)	T2 Initial	T2 Final	T2 Diff. (mL)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	AnalDate	Anal Time
1	PB128592BL		1	50	50	5.00	0.00	4.82	4.82	0.18	0.00	0.00	04/30/2020	13:12
2	PB128592BS	25	1	50	50	5.00	0.00	2.22	2.22	2.78	2.60	20.80	04/30/2020	13:15
3	L2182-12		1	50	50	5.00	0.00	4.78	4.78	0.22	0.04	0.32	04/30/2020	13:18

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume