



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB108956

Slope : 97.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/30/2020	17:10
2	CAL2	1	Water	NA	NA	20.2	7.00	04/30/2020	17:11
3	CAL3	1	Water	NA	NA	20.2	10.02	04/30/2020	17:12
4	ICV	1	Water	NA	NA	20.2	6.98	04/30/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.02	04/30/2020	17:16
6	L2122-15	1	Solid	20.03	20	21.5	8.85	04/30/2020	17:19
7	L2122-15DUP	1	Solid	20.04	20	21.6	8.86	04/30/2020	17:20
8	L2452-02	1	Solid	20.02	20	22.5	6.35	04/30/2020	17:21
9	L2452-04	1	Solid	20.03	20	22.9	7.68	04/30/2020	17:22
10	L2452-06	1	Solid	20.04	20	22.6	7.64	04/30/2020	17:24
11	L2452-08	1	Solid	20.03	20	22.4	7.70	04/30/2020	17:25
12	L2463-06	1	Solid	20.02	20	21.5	8.08	04/30/2020	17:26
13	L2463-12	1	Solid	20.03	20	21.6	8.19	04/30/2020	17:27
14	L2463-15	1	Solid	20.05	20	21.9	8.30	04/30/2020	17:29
15	L2463-19	1	Solid	20.03	20	21.8	8.86	04/30/2020	17:30
16	CCV2	1	Water	NA	NA	20.1	12.01	04/30/2020	17:32
17	L2463-22	1	Solid	20.04	20	21.8	7.75	04/30/2020	17:33
18	CCV3	1	Water	NA	NA	20.1	2.01	04/30/2020	17:34

108956

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2452 WorkList ID : 138351 Department : Wet-Chemistry Date : 04-30-2020 11:38:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2020	Solid	L2122-15	Corrosivity	Cool 4 deg C	PSEG05	L13	SU-03-043020	04/30/2020	9045D
04/29/2020	Solid	L2452-02	Corrosivity	Cool 4 deg C	PSEG01	L13	72-11976	04/29/2020	9045D
04/29/2020	Solid	L2452-04	Corrosivity	Cool 4 deg C	PSEG01	L13	224	04/29/2020	9045D
04/29/2020	Solid	L2452-06	Corrosivity	Cool 4 deg C	PSEG01	L13	227	04/29/2020	9045D
04/29/2020	Solid	L2452-08	Corrosivity	Cool 4 deg C	PSEG01	L13	209	04/29/2020	9045D
04/30/2020	Solid	L2463-06	Corrosivity	Cool 4 deg C	TULL02	K31	AOC14-CARWASH	04/30/2020	9045D
04/30/2020	Solid	L2463-12	Corrosivity	Cool 4 deg C	TULL02	K31	NORTH-AOC14-CAR-WASH	04/30/2020	9045D
04/30/2020	Solid	L2463-15	Corrosivity	Cool 4 deg C	TULL02	K31	AOC3-FORM-U-SUPPLEMEN	04/30/2020	9045D
04/30/2020	Solid	L2463-19	Corrosivity	Cool 4 deg C	TULL02		JHDS-SOUTH-D	04/30/2020	9045D
04/30/2020	Solid	L2463-22	Corrosivity	Cool 4 deg C	TULL02		JHDS-SOUTH-C	04/30/2020	9045D

Date/Time 04/30/2020 15:40
 Received by: JP
 Relinquished by: JP

Date/Time 04/30/2020 17:35
 Received by: JP
 Relinquished by: JP