



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108960

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	04/30/2020	16:30
2	CAL2	1	Water	NA	NA	20.2	7.01	04/30/2020	16:32
3	CAL3	1	Water	NA	NA	20.2	10.03	04/30/2020	16:33
4	ICV	1	Water	NA	NA	20.2	6.99	04/30/2020	16:35
5	CCV1	1	Water	NA	NA	20.1	2.02	04/30/2020	16:36
6	L2456-03	1	Water	NA	NA	20.6	7.52	04/30/2020	16:37
7	L2457-03	1	Water	NA	NA	20.7	7.14	04/30/2020	16:40
8	L2457-04	1	Water	NA	NA	20.6	7.17	04/30/2020	16:42
9	L2462-01	1	Water	NA	NA	15.8	6.94	04/30/2020	16:45
10	L2462-01DUP	1	Water	NA	NA	15.9	6.95	04/30/2020	16:46
11	CCV2	1	Water	NA	NA	20.1	12.02	04/30/2020	16:48

WORKLIST(Hardcopy Internal Chain)

1308960

WorkList Name : PHWL2462 WorkList ID : 138358 Department : Wet-Chemistry Date : 04-30-2020 15:52:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
04/30/2020	Water	L2456-03	pH - H	Cool 4 deg C	PSEG01	L22	200	04/30/2020	9040C
04/30/2020	Water	L2457-03	pH - H	Cool 4 deg C	PSEG01	L21	1342	04/30/2020	9040C
04/30/2020	Water	L2457-04	pH - H	Cool 4 deg C	PSEG01	L21	1341	04/30/2020	9040C
04/30/2020	Water	L2462-01	pH - F	Cool 4 deg C	TULL02	K22	AOC-14	04/30/2020	9040C

Date/Time 04/30/2020 16:00
 Received by: Jn
 Relinquished by: Jn

Date/Time 04/30/2020 18:00
 Received by: Jn
 Relinquished by: Jn