

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 05/02/2020

OVEN IN TEMP1: 98 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 05/01/2020 16:00

DATE & TIME OUT1: 05/02/2020 07:10

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 05/02/2020 08:00

DATE & TIME OUT2: 05/02/2020 09:30

Run Number: LB108965

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB108965BL	LB108965BL	78.5687	78.5687	100	78.5687	78.5687	78.5687	0.0000	0
2	L2463-16	AOC3-FORM-U-SUPPLEMENTAL	74.7982	74.7982	100	74.8217	74.8217	74.8217	0.0235	235
3	L2463-16DUP	AOC3-FORM-U-SUPPLEMENTALDUP	68.8148	68.8148	100	68.8391	68.8391	68.8391	0.0243	243

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

LB108965

WorkList Name : ASTM TS L2463 WorkList ID : 138386 Department : Wet-Chemistry Date : 05-01-2020 11:17:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2020	Solid	L2463-16	ASTM TS	Cool 4 deg C	TULL02	K31	AOC3-FORM-U-SUPPLEMEN	04/30/2020	SM2540 B

Date/Time 05.01.2020 15:00
 Received by: JA
 Relinquished by: KS

Date/Time 05.01.2020 17:00
 Received by: KS
 Relinquished by: JA