


LB108967

245.1

INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.079	-0.079	633	0.000	633				
0.05	0.050	0.049	-0.001	1123	0.0 %	1123				
0.2	0.200	0.229	0.029	1811	0.0 %	1811				
2.5	2.500	2.528	0.028	10588	0.0 %	10588				
5.0	5.000	5.022	0.022	20113	0.0 %	20113				
7.5	7.500	7.588	0.088	29913	0.0 %	29913				
10.0	10.000	9.913	-0.087	38792	0.0 %	38792				

LB108967 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 May 2020 11:41:10

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	633	-	0.0000	245.1	PPB	01 May 2020 11:43:21	S	Std
0.05 - 1	50.05	1123	-	0.0500	245.1	PPB	01 May 2020 11:45:38	S	Std
0.2 - 1	50.2	1811	-	0.2000	245.1	PPB	01 May 2020 11:47:55	S	Std
2.5 - 1	52.5	10588	-	2.5000	245.1	PPB	01 May 2020 11:50:12	S	Std
5.0 - 1	55	20113	-	5.0000	245.1	PPB	01 May 2020 11:52:28	S	Std
7.5 - 1	57.5	29913	-	7.5000	245.1	PPB	01 May 2020 11:54:44	S	Std
10.0 - 1	510	38792	-	10.0000	245.1	PPB	01 May 2020 11:59:40	S	Std
ICV26 - 1	ICV26	16678	4.1224	-245.1		PPB	01 May 2020 12:04:57	U	SMPL
ICB26 - 1	ICB26	634	-0.0789	-245.1		PPB	01 May 2020 12:09:45	U	SMPL
CCV31 - 1	CCV31	20353	5.0847	-245.1		PPB	01 May 2020 12:12:00	U	SMPL
CCB31 - 1	CCB31	696	-0.0627	-245.1		PPB	01 May 2020 12:16:47	U	SMPL
CRA - 1	CRA	1115	0.0470	-245.1		PPB	01 May 2020 12:19:02	U	SMPL
HighStd - 1	HighStd	38650	9.8760	-245.1		PPB	01 May 2020 12:21:17	U	SMPL
ChkStd - 1	ChkStd	28607	7.2461	-245.1		PPB	01 May 2020 12:23:31	U	SMPL
PB128604BL - 1	PBW	664	-0.0711	-245.1		PPB	01 May 2020 12:30:58	U	SMPL
PB128604BS - 1	LCSW	16084	3.9668	-245.1		PPB	01 May 2020 12:35:57	U	SMPL
L2462-01 - 1	AOC-14	2938	0.5244	-245.1		PPB	01 May 2020 12:40:47	U	SMPL
L2462-01DUP - 1	AOC-14DUP	3284	0.6150	-245.1		PPB	01 May 2020 12:43:03	U	SMPL
L2462-01MS - 1	AOC-14MS	18589	4.6228	-245.1		PPB	01 May 2020 12:45:20	U	SMPL
L2462-01MSD - 1	AOC-14MSD	18569	4.6175	-245.1		PPB	01 May 2020 12:47:37	U	SMPL
L2462-02 - 1	AOC-14	769	-0.0436	-245.1		PPB	01 May 2020 12:52:23	U	SMPL
CCV32 - 1	CCV32	20418	5.1017	-245.1		PPB	01 May 2020 12:54:38	U	SMPL
CCB32 - 1	CCB32	543	-0.1027	-245.1		PPB	01 May 2020 12:56:53	U	SMPL
L2462-01LX5 - 1		1139	0.0533	-245.1		PPB	01 May 2020 12:59:08	U	SMPL
L2462-01A - 1		17899	4.4421	-245.1		PPB	01 May 2020 13:01:23	U	SMPL
CCV33 - 1	CCV33	19855	4.9543	-245.1		PPB	01 May 2020 13:03:39	U	SMPL
CCB33 - 1	CCB33	593	-0.0896	-245.1		PPB	01 May 2020 13:05:54	U	SMPL