



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB108975

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/01/2020	10:45
2	L2456-03		1	8	100.00	>212.0	>212.0	05/01/2020	11:20
3	L2456-03DUP		1	8	100.00	>212.0	>212.0	05/01/2020	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp5120

WorkList ID : 138381

Department : Wet-Chemistry

Date : 05-01-2020 09:55:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/07/2020	Water	L2456-03	Flash Point	Cool 4 deg C	PSEG01	L22	200	04/30/2020	1010A

Date/Time

05/01/20 10:15

Received by:

KS

Relinquished by:

KS

Date/Time

05/01/20 11:55

Received by:

KS

Relinquished by:

KS