



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB108978

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP83253
COD calibration std. 100 ppm	WP83249
COD calibration std. 10 ppm	WP83247
COD calibration std. 150 ppm	WP83250
COD calibration std. 50 ppm	WP83252
COD calibration std. 0 ppm	WP83246
COD Digestion Vials Low Level 0-150Mg/L	W2678
COD ICV-LCS std, 50ppm	WP83913
COD CCV std, 50ppm	WP83912
COD LOD std, 5ppm	WP83929
COD Digestion Vials Low Level 0-150Mg/L	W2679

Temp In(C): 149	Date In: 05/01/2020	Time In: 12:10
Temp Out(C): 150	Date Out: 05/01/2020	Time Out: 14:10

Intercept: 0.4144

Slope: 1.0094

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	04/03/2020	18:04
2	CAL2	10	1	Water	11.000	04/03/2020	18:05
3	CAL3	50	1	Water	51.000	04/03/2020	18:05
4	CAL4	100	1	Water	101.000	04/03/2020	18:06
5	CAL5	150	1	Water	152.000	04/03/2020	18:06



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB108978

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	49.000	48.133	04/03/2020	18:07
2	ICB				1	Water	0.000	-0.411	04/03/2020	18:07
3	CCV1	50			1	Water	50.000	49.124	05/01/2020	15:59
4	CCB1				1	Water	0.000	-0.411	05/01/2020	16:00
5	LB108978BL		5.00	50	1	Solid	0.000	-0.411	05/01/2020	16:00
6	LB108978BS	50	5.00	50	1	Solid	49.000	48.133	05/01/2020	16:01
7	L2182-03		5.01	50	1	Solid	4.000	3.552	05/01/2020	16:01
8	CCV2	50			1	Water	50.000	49.124	05/01/2020	16:02
9	CCB2				1	Water	0.000	-0.411	05/01/2020	16:02