



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108988

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/01/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	05/01/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	05/01/2020	17:12
4	ICV	1	Water	NA	NA	20.1	6.99	05/01/2020	17:15
5	CCV1	1	Water	NA	NA	20.2	2.01	05/01/2020	17:25
6	L2464-03	1	Solid	20.03	20	22.5	6.10	05/01/2020	17:26
7	L2464-03DUP	1	Solid	20.04	20	22.6	6.11	05/01/2020	17:27
8	L2464-04	1	Solid	20.02	20	22.7	6.57	05/01/2020	17:30
9	CCV2	1	Water	NA	NA	20.2	12.02	05/01/2020	17:32

LB 108988

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH SOLIDS L2464 WorkList ID : 138407 Department : Wet-Chemistry Date : 05-01-2020 16:58:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2020	Solid	L2464-03 ✓	pH	Cool 4 deg C	PSEG01	L23	TRE-50006-LITTER	05/01/2020	9045D
05/01/2020	Solid	L2464-04 ✓	pH	Cool 4 deg C	PSEG01	L23	TRE-50006-SOIL	05/01/2020	9045D

Date/Time 05.01.2020 17:00
 Received by: JP
 Relinquished by: JP

Date/Time 05.01.2020 18:10
 Received by: JP
 Relinquished by: JP