



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB108993

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/04/2020	11:40
2	CAL2	1	Water	NA	NA	20.1	7.00	05/04/2020	11:41
3	CAL3	1	Water	NA	NA	20.1	10.03	05/04/2020	11:42
4	ICV	1	Water	NA	NA	20.2	6.98	05/04/2020	11:45
5	CCV1	1	Water	NA	NA	20.1	2.02	05/04/2020	11:46
6	L2485-01	1	Water	NA	NA	21.6	5.81	05/04/2020	11:48
7	L2485-01DUP	1	Water	NA	NA	21.7	5.82	05/04/2020	11:49
8	CCV2	1	Water	NA	NA	20.1	12.02	05/04/2020	11:50

LB 108993

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2485 WorkList ID : 138431 Department : Wet-Chemistry Date : 05-04-2020 11:31:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/01/2020	Water	L2485-01 - ☒	pH	Cool 4 deg C	STER01	K22	POOL-WATER	05/01/2020	9040C

Date/Time 05/04/2020 11:35
Received by: Jo
Relinquished by: A/n

Date/Time 05/04/2020 12:45
Received by: A/n
Relinquished by: Jo