


LB108999

7470A

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.5164e-004

C= 1.7836e-002

Rho= 0.9997226

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	54	0.000	54				
0.2	0.200	0.155	-0.045	544	0.0 %	544				
2.5	2.500	2.458	-0.042	9698	0.0 %	9698				
5.0	5.000	5.159	0.159	20430	0.0 %	20430				
7.5	7.500	7.378	-0.122	29248	0.0 %	29248				
10.0	10.000	10.019	0.019	39745	0.0 %	39745				

LB108999 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 May 2020 13:38:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7470A	PPB	04 May 2020 13:39:59	S	Std
0.2 - 1	50.2	544	-	0.2000	7470A	PPB	04 May 2020 13:42:06	S	Std
2.5 - 1	52.5	9698	-	2.5000	7470A	PPB	04 May 2020 13:44:14	S	Std
5.0 - 1	55	20430	-	5.0000	7470A	PPB	04 May 2020 13:46:21	S	Std
7.5 - 1	57.5	29248	-	7.5000	7470A	PPB	04 May 2020 13:48:28	S	Std
10.0 - 1	510	39745	-	10.0000	7470A	PPB	04 May 2020 13:50:35	S	Std
ICV30 - 1	ICV30	15861	4.0090	-	7470A	PPB	04 May 2020 13:55:35	U	SMPL
ICB30 - 1	ICB30	-82	-0.0028	-	7470A	PPB	04 May 2020 13:57:40	U	SMPL
CCV45 - 1	CCV45	20273	5.1193	-	7470A	PPB	04 May 2020 13:59:46	U	SMPL
CCB45 - 1	CCB45	-98	-0.0068	-	7470A	PPB	04 May 2020 14:01:51	U	SMPL
CRA - 1	CRA	512	0.1467	-	7470A	PPB	04 May 2020 14:03:58	U	SMPL
HighStd - 1	HighStd	39157	9.8712	-	7470A	PPB	04 May 2020 14:06:03	U	SMPL
ChkStd - 1	ChkStd	26999	6.8118	-	7470A	PPB	04 May 2020 14:08:09	U	SMPL
PB128639BL - 1	PBW	-162	-0.0229	-	7470A	PPB	04 May 2020 14:10:17	U	SMPL
PB128639BS - 1	LCSW	14810	3.7446	-	7470A	PPB	04 May 2020 14:12:26	U	SMPL
L2457-03 - 1	1342	-102	-0.0078	-	7470A	PPB	04 May 2020 14:14:33	U	SMPL
L2457-03DUP - 1	1342DUP	0	0.0178	-	7470A	PPB	04 May 2020 14:16:40	U	SMPL
L2457-03MS - 1	1342MS	14937	3.7765	-	7470A	PPB	04 May 2020 14:18:48	U	SMPL
L2457-03MSD - 1	1342MSD	16181	4.0896	-	7470A	PPB	04 May 2020 14:23:24	U	SMPL
L2457-04 - 1	1341	-96	-0.0063	-	7470A	PPB	04 May 2020 14:25:29	U	SMPL
CCV46 - 1	CCV46	20440	5.1613	-	7470A	PPB	04 May 2020 14:27:35	U	SMPL
CCB46 - 1	CCB46	-105	-0.0086	-	7470A	PPB	04 May 2020 14:29:41	U	SMPL
L2464-07 - 1	TRE-50029	69	0.0352	-	7470A	PPB	04 May 2020 14:31:48	U	SMPL
PB128639TB - 1	PB128627TB	-10	0.0153	-	7470A	PPB	04 May 2020 14:33:54	U	SMPL
L2457-03LX5 - 1		10	0.0204	-	7470A	PPB	04 May 2020 14:35:59	U	SMPL
L2457-03A - 1		15513	3.9215	-	7470A	PPB	04 May 2020 14:38:06	U	SMPL
CCV47 - 1	CCV47	20061	5.0659	-	7470A	PPB	04 May 2020 14:40:12	U	SMPL
CCB47 - 1	CCB47	-106	-0.0088	-	7470A	PPB	04 May 2020 14:42:19	U	SMPL