



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109003

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP84025
Turbidity Calibration std, 0NTU	WP84024
Turbidity Calibration std, 20NTU	WP84019
Turbidity Calibration std, 40NTU	WP84018
Turbidity Calibration std, 1NTU	WP84022
Turbidity Calibration std, 5NTU	WP84021
Turbidity LOQ std, 1.0NTU	WP84026
Turbidity Calibration - CCV std, 10 NTU	WP84020
10 NTU Standard 500 ml	W2489

Intercept: 0.1939

Slope: 1.1225

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.275	05/04/2020	14:22
2	CAL2	1	1	1.320	05/04/2020	14:25
3	CAL3	5	1	5.980	05/04/2020	14:28
4	CAL4	10	1	11.200	05/04/2020	14:32
5	CAL5	20	1	22.500	05/04/2020	14:35
6	CAL6	40	1	45.200	05/04/2020	14:38



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109003

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.700	9.360	05/04/2020	14:42
2	ICB		1	0.258	0.057	05/04/2020	14:45
3	CCV1	10	1	10.900	9.538	05/04/2020	14:48
4	CCB1		1	0.266	0.064	05/04/2020	14:51
5	LB109003BL		1	0.272	0.070	05/04/2020	14:55
6	L2182-10		1	0.722	0.470	05/04/2020	14:58
7	L2182-11		1	1.310	0.994	05/04/2020	15:01
8	CCV2	10	1	10.800	9.449	05/04/2020	15:04
9	CCB2		1	0.254	0.054	05/04/2020	15:08